

RECEIVED

By Tracy Crews at 2:37 pm, Nov 09, 2020



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005854	LOCATION OF INSTRUMENT LIBERTY POLICE DEPT.	DATE OF INSPECTION 08/12/2020	TIME OF INSPECTION 07:00
---------------------------------------	--	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE DRY	STANDARD LOT # AG924701	STANDARD EXPIRATION DATE 09/04/2021
-----	-----	-----	SIM TEMPERATURE N/A	SIM SERIAL NUMBER N/A	SIM CERTIFICATE EXPIRATION N/A
Air Blank	0.000	07:02	STANDARD VALUE 0.100	STANDARD SUPPLIER INTOXIMETERS	
Cal Check	0.100	07:02	CALIBRATION CHECK RESULT 1 0.100		
Air Blank	0.000	07:03	CALIBRATION CHECK RESULT 2 0.101		
Cal Check	0.101	07:03	CALIBRATION CHECK RESULT 3 0.100		
Air Blank	0.000	07:04	MAXIMUM DEVIATION (MUST BE WITHIN 5%) 1.0%		
Cal Check	0.100	07:04	SPREAD (MUST BE .005 OR LESS) 0.001		
Air Blank	0.000	07:05			
Pass					

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Voltage/Current Test	Pass		Test	g/210L	Time
RAM Test	Pass		-----	-----	-----
EEPROM Checksum Test	Pass		Air Blank	0.000	07:05
Real Time Clock Test	Pass		Subject Test	RFI*	07:05
DSP Test	Pass		Air Blank	0.000	07:06
Analytical Stability Test	Pass		*RFI Detect		
Modem Test	Pass				
Temperature Regulation Test	Pass				
Pass			Pass		

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT						
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19	
0	0	0	0	0	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

OPERATING WITHIN SPECIFICATIONS

INSPECTING OFFICER

SIGNATURE 	PRINT NAME ROBERT D BRATCHER
---------------	---------------------------------

TYPE II PERMIT NUMBER 290182	EXPIRATION DATE 08/19/2021	TELEPHONE NUMBER 8164394701
---------------------------------	-------------------------------	--------------------------------