



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005854	LOCATION OF INSTRUMENT LIBERTY POLICE DEPT.	DATE OF INSPECTION 12/01/2020	TIME OF INSPECTION 10:41
--	---	---	------------------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
---	---	---	DRY	AG924701	09/04/2021
Air Blank	0.000	10:44	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Cal Check	0.102	10:44	N/A	N/A	N/A
Air Blank	0.000	10:45	STANDARD VALUE	STANDARD SUPPLIER	
Cal Check	0.101	10:45	0.100	INTOXIMETERS	
Air Blank	0.000	10:45	CALIBRATION CHECK RESULT 1		
Cal Check	0.102	10:46	0.102		
Air Blank	0.000	10:46	CALIBRATION CHECK RESULT 2		
Cal Check	0.101	10:46	0.101		
Air Blank	0.000	10:46	CALIBRATION CHECK RESULT 3		
			0.102		
			MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)	
			2.0%	0.001	

Pass

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
			Test	g/210L	Time
Voltage/Current Test	Pass		---	---	---
RAM Test	Pass		Air Blank	0.000	10:47
EEPROM Checksum Test	Pass		Subject Test	RFI*	10:47
Real Time Clock Test	Pass		Air Blank	0.000	10:48
DSP Test	Pass		*RFI Detect		
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				

Pass

Pass

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT						
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19	
0	0	0	0	0	0	1

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).
OPERATING WITHIN SPECIFICATIONS

INSPECTING OFFICER		
SIGNATURE 	PRINT NAME ROBERT D BRATCHER	
PERMIT NUMBER 290182	EXPIRATION DATE 08/19/2021	TELEPHONE NUMBER 8164394701

