



ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
By Tracy Crews at 10:13 am, Mar 27, 2021 RT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 030788	NAME OF AGENCY Jefferson County Sheriff's Office	DATE OF INSPECTION 03/10/2021
LOCATION OF INSTRUMENT (STREET AND CITY) 410 First Street, Hillsboro, MO 63050		TIME OF INSPECTION 7:30 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

<input type="checkbox"/> SIMULATOR SOLUTION	<input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>Intoximeters</u>	LOT # <u>AG912204</u> EXP. DATE <u>05/02/2021</u>
<input type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C)	SIM. SN _____ SIM. NIST EXP DATE _____

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 → .095	TEST 2 → .095	TEST 3 → .095
---------------	---------------	---------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	---------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE 	PRINT NAME Deputy Shawn Lones #391
TYPE II PERMIT NUMBER/EXPIRATION DATE 200300 12/11/2022	TELEPHONE NUMBER (636) 797-5000

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 030708
Version no: 532B

TEST RECORD 00407

Temp Date Time 210L

Air Blank:
03/10/21 18:28 .000
Calibration Check:
24 03/10/21 18:28 .095

Subject Name
MARCH 2021

Subject I.D.
M41

Operator Name, I.D.

Location
TEST #1

AS IV Serial no: 030708
Version no: 532B

TEST RECORD 00408

Temp Date Time 210L

Air Blank:
03/10/21 18:30 .000
Calibration Check:
24 03/10/21 18:30 .095

Subject Name

Subject I.D.

Operator Name, I.D.

Location
TEST #2

AS IV Serial no: 030708
Version no: 532B

TEST RECORD 00409

Temp Date Time 210L

Air Blank:
03/10/21 18:32 .000
Calibration Check:
25 03/10/21 18:32 .095

Subject Name

Subject I.D.

Operator Name, I.D.

Location
TEST #3

AS IV Serial no: 030708
Version no: 532B

TEST RECORD 00410

Temp Date Time 210L

VOID: RFI
12 03/10/21 18:34

Subject Name

Subject I.D.

Operator Name, I.D.

Location
RFI!



Certificate of Analysis

Customer Name
Exclusive Supplier
Proximeters, Inc
2081 Craig Road
St Louis, Mo 63140

Test Date: 6-May-2019

Lot # AG912204 Model 108cacc

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
11/13/19	Ice	Ethanol Nitrogen	0.100 ± 2% BAC (272 ppm) Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm
<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC434668	800.0 ppm	0056649	390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

Analytical Method: NDIR

Approved for Release:

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07