



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
 By Tracy Crews at 2:33 pm, Dec 01, 2021

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 03079	NAME OF AGENCY Hazelwood Police Department	DATE OF INSPECTION 11/30/2021
----------------------------	---	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 415 Elm Grove Ln. / Hazelwood, MO 63042	TIME OF INSPECTION 3:30 pm
---	-------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

<input checked="" type="checkbox"/> SIMULATOR SOLUTION	<input type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>Guth Laboratories, INC</u>	LOT # <u>21080</u> EXP. DATE <u>03/08/2023</u>
<input checked="" type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C) <u>34.0</u>	SIM. SN <u>MP6029</u> SIM. NIST EXP DATE <u>07/19/2022</u>

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 \bullet .100	TEST 2 \bullet .104	TEST 3 \bullet .103
-----------------------	-----------------------	-----------------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

BA Maint. 11/30/2021
 Simulator Solution Bottle Number 709

 Changed the time due to Daylight Savings

INSPECTING OFFICER	
SIGNATURE 	PRINT NAME Jason Corson
TYPE II PERMIT NUMBER/EXPIRATION DATE 210057 / 4/8/2023	TELEPHONE NUMBER (314) 838-5000

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 030791
Version no: 532B

TEST RECORD 00732 ✓

Temp Date Time 210L

Air Blank: 11/30/21 15:35 .000
Calibration Check: 20 11/30/21 15:35 .100

Subject Name

Test # 1

Subject I.D.

Corson 210057

Operator Name, I.D.

Hazelwood BAT

Location

Van

AS IV Serial no: 030791
Version no: 532B

TEST RECORD 00733 ✓

Temp Date Time 210L

Air Blank: 11/30/21 15:37 .000
Calibration Check: 22 11/30/21 15:37 .104

Subject Name

Test # 2

Subject I.D.

Corson 210057

Operator Name, I.D.

Hazelwood BAT

Location

Van

AS IV Serial no: 030791
Version no: 532B

TEST RECORD 00734 ✓

Temp Date Time 210L

Air Blank: 11/30/21 15:39 .000
Calibration Check: 23 11/30/21 15:39 .103

Subject Name

Test # 3

Subject I.D.

Corson 210057

Operator Name, I.D.

Hazelwood BAT

Location

Van

AS IV Serial no: 030791
Version no: 532B

TEST RECORD 00735 ✓

Temp Date Time 210L

VOID: RFI
12 11/30/21 15:40

Subject Name

Test # 4

Subject I.D.

Corson 210057

Operator Name, I.D.

RFI Test

Location

Hazelwood BAT

Van

AS IV Serial no: 030791
Version no: 532B

TEST RECORD 00736 ✓

Temp Date Time 210L

Air Blank: 11/30/21 15:45 .000
Subject Test: Auto 24 11/30/21 15:45 .000

Subject Name

Test # 5

Subject I.D.

Corson 210057

Operator Name, I.D.

Blank Test

Location

Hazelwood BAT

Van



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **21080** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **March 10, 2021**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1214%** (w/vol) ethyl alcohol. The expiration date for this lot number is **March 8, 2023** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03052002 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



Missouri Department of Health and Senior Services
 P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010
 RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466
Robert J. Knodell
 Acting Director



Michael L. Parson
 Governor

SIMULATOR CERTIFICATION REPORT

SIMULATOR INFORMATION

Simulator Serial Number: MP6029 **Manufacturer:** Guth
Model Number: 12V500
Agency: HAZELWOOD PD
Agency Address: 415 ELM GROVE LANE, HAZELWOOD, MO 63042

NIST THERMOMETER INFORMATION

Serial Number: 17KMM00690 **Bias:** 0.00
Uncertainty: 0.02
Date of Certification: 11/6/2020 **Date of Expiration:** 11/6/2021

ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

VERIFICATION RESULTS

<u>Simulator Average</u>	<u>NIST Average</u>	<u>Combined Uncertainty</u>
34.00	34.01	.03

The combined uncertainty is calculated with a k=2 value.

ADJUSTMENT RESULTS

No adjustment was needed.

Date of testing: 7/19/2021
Certification Expiration: 7/19/2022
Simulator testing technician: D. DEBOARD

Notes on Condition: none

Deviation(s) from method: none

DHSS BAP Scientist Approving: B. LUTMER
Certification No: MP6029_7192021

X

DHSS BAP Scientist Approving



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
JASON CORSON

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/6/2021

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 210057

EXPIRES 4/6/2023

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 600-0771 (6-10)

LAB-4 (P6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator CORSON, JASON
Permit No 210057
Date Issued 4/6/2021 Date Expires 4/6/2023