

AS IV Serial no: 035484
Version no: 532C

TEST RECORD 00161

Temp Date Time 210L ^{s/}

Air Blank:
11/06/21 10:53 .000
Calibration Check:
26 11/06/21 10:53 .083

Subject Name

Test #1

Subject I.D.

Operator Name, I.D.

Cpl. Johnson 200148

Location

SCCPD

AS IV Serial no: 035484
Version no: 532C

TEST RECORD 00162

Temp Date Time 210L ^{s/}

Air Blank:
11/06/21 10:55 .000
Calibration Check:
26 11/06/21 10:55 .082

Subject Name

Test #2

Subject I.D.

Operator Name, I.D.

Cpl. Johnson 200148

Location

SCCPD

AS IV Serial no: 035484
Version no: 532C

TEST RECORD 00163

Temp Date Time 210L ^{s/}

Air Blank:
11/06/21 10:56 .000
Calibration Check:
27 11/06/21 10:56 .082

Subject Name

Test #3

Subject I.D.

Operator Name, I.D.

Cpl. Johnson 200148

Location

SCCPD

AS IV Serial no: 035484
Version no: 532C

TEST RECORD 00164

Temp Date Time 210L ^{s/}

VOID: RFI
12 11/06/21 10:57

Subject Name

KFI Test

Subject I.D.

Operator Name, I.D.

Cpl. Johnson 200148

Location

SCCPD

AS IV Serial no: 035484
Version no: 532C

TEST RECORD 00165

Temp Date Time 210L ^{s/}

Air Blank:
11/06/21 10:59 .000
Subject Test: Auto
27 11/06/21 10:59 .000

Subject Name

Self-Test

Subject I.D.

Cpl. Johnson 200148

Operator Name, I.D.

Location

SCCPD



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 7-Apr-2021

Lot # AG109701 Model 108cacd

<u>Exp. Date</u> 7-Apr-2023	<u>Cyl. Type</u> 108	<u>Component</u> Ethanol Nitrogen	<u>Certified Concentration</u> 0.082 ± 0.002 BrAC (223 ppm) Balance
---------------------------------------	--------------------------------	--	--

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u> EB0010581 EB0010570 EB0010285 EB0010561 EB0010681	<u>Concentration</u> 392.1 ppm 259.8 ppm 208.0 ppm 103.6 ppm 52.12 ppm	<u>RGM Serial No.</u> EB0010603 EB0010589 EB0010596 EB0010582 EB0010579	<u>Concentration</u> 393.0 ppm 258.2 ppm 208.3 ppm 104.2 ppm 52.81 ppm
<u>CRM Serial No.</u> CC727481 CC727496	<u>Concentration</u> 800.0 ppm 253.0 ppm	<u>CRM Serial No.</u> CC727493 CC727498	<u>Concentration</u> 390.0 ppm 150.0 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
 Date: 2021.04.07 17:42:01 -05:00
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)

Approved for Release:

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
MIKE JOHNSON

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/8/2020

NUMBER 200198

EXPIRES 7/8/2022

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)

 STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator JOHNSON, MIKE
 Permit No 200198
 Date Issued 7/8/2020 Date Expires 7/8/2022

