



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 062087	PRINTER SN 03A.2436.034	DATE OF INSPECTION 02/03/2021
LOCATION OF INSTRUMENT (STREET AND CITY) 4001 NE Lakewood Court, Lee's Summit, MO 64064		TIME OF INSPECTION 0:53 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER Intoximeters LOT # AG006306 EXP. DATE 03/03/2022
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIMULATOR SN _____ SIMULATOR EXP DATE _____

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ➡ .102	TEST 2 ➡ .103	TEST 3 ➡ .102
---------------	---------------	---------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).
 Instrument meets all DHSS standards and guidelines.

INSPECTING OFFICER	
SIGNATURE 	PRINT NAME Dep. D.B. LOVE #61
TYPE II PERMIT NUMBER/EXPIRATION DATE 200052/ 01-10-2022	TELEPHONE NUMBER (816) 795-1960

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901

AS IV Serial no: 062087
Version no: 532B

TEST RECORD 00962

Temp Date Time 210L ^{g/}

Air Blank:

02/03/21 00:53 .000

Calibration Check:
21 02/03/21 00:53 .102

Subject Name

Monthly Maint

Subject I.D.

Operator Name: I.D.

D.B. Love #61

Location

JCSO - GHQ

DHSS 20052

AS IV Serial no: 062087
Version no: 532B

TEST RECORD - REPRINT

TEST RECORD 00961

Temp Date Time 210L ^{g/}

Air Blank:

02/03/21 00:50 .000

Calibration Check:
20 02/03/21 00:50 .103

Subject Name

Monthly Maint

Subject I.D.

Operator Name: I.D.

D.B. Love #61

Location

JCSO - GHQ

DHSS-20052

AS IV Serial no: 062087
Version no: 532B

TEST RECORD 00963

Temp Date Time 210L ^{g/}

Air Blank:

02/03/21 00:56 .000

Calibration Check:
23 02/03/21 00:56 .102

Subject Name

Monthly Maint

Subject I.D.

Operator Name: I.D.

D.B. Love #61

Location

JCSO - GHQ

DHSS-20052

AS IV Serial no: 062087
Version no: 532B

TEST RECORD - REPRINT

TEST RECORD 00964

Temp Date Time 210L ^{g/}

VOID: RFI

12 02/03/21 00:59

Subject Name

Monthly Maint

Subject I.D.

Operator Name: I.D.

D.B. Love #61

Location

JCSO - GHQ

DHSS-20052



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
DUSTIN B. LOVE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 1/10/2020

NUMBER 200052

EXPIRES 1/10/2022

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

 **STATE OF MISSOURI**
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator LOVE, DUSTIN
Permit No 200052
Date Issued 1/10/2020 **Date Expires** 1/10/2022

