



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 062093	NAME OF AGENCY Sugar Creek Police Department	DATE OF INSPECTION 01/23/2021
LOCATION OF INSTRUMENT (STREET AND CITY) 1001 Heroes Way, Sugar Creek, MO		TIME OF INSPECTION 5:09 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER RepCo LOT # 19002 EXP. DATE 10/16/2021

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIM. SN SD 3136 SIM. NIST EXP DATE 08/12/2021

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .103

TEST 2 .103

TEST 3 .103

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Replaced printer battery at MO Safety Center Lab (Warrensburg) 1-19-21

INSPECTING OFFICER

SIGNATURE

PRINT NAME
Garrett Holman

TYPE II PERMIT NUMBER/EXPIRATION DATE
200126 03/09/2022

TELEPHONE NUMBER
(816) 252-5560

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

SUGAR CREEK MISSOURI POLICE DEPARTMENT

ALCO-SENSOR IV WITH PRINTER TEST RESULTS

TEST # 1

AS IV Serial no: 062093
Version no: 532B

TEST RECORD 01289
Temp Date Time 210L

Air Blank:
01/23/21 17:09 .000
Calibration Check:
24 01/23/21 17:09 .103

Subject Name
TEST # 1

Subject I.D.
Holman
Operator Name, I.D.

Location
SCPD

TEST # 2

AS IV Serial no: 062093
Version no: 532B

TEST RECORD 01290
Temp Date Time 210L

Air Blank:
01/23/21 17:11 .000
Calibration Check:
25 01/23/21 17:11 .103

Subject Name
TEST # 2

Subject I.D.
Holman
Operator Name, I.D.

Location
SCPD

TEST # 3

AS IV Serial no: 062093
Version no: 532B

TEST RECORD 01291
Temp Date Time 210L

Air Blank:
01/23/21 17:13 .000
Calibration Check:
25 01/23/21 17:13 .103

Subject Name
TEST # 3

Subject I.D.
Holman
Operator Name, I.D.

Location
SCPD

RFI TEST

AS IV Serial no: 062093
Version no: 532B

TEST RECORD 01292
Temp Date Time 210L

VOID: RFI
12 01/23/21 17:14

Subject Name
RFI TEST

Subject I.D.
Holman
Operator Name, I.D.

Location
SCPD

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing Co.

LOT NUMBER: 19002

EXPIRATION DATE: October 16, 2021 at 11:59 p.m.

RepCo Marketing Co. certifies the following:

RepCo Marketing Co. prepared, tested and supplied Lot Number 19002 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by ISO 17025 and ISO 17034 accredited institution Alcohol Countermeasure Systems, using NIST standards. Random samples were analyzed by Alcohol Countermeasure Systems utilizing a gas chromatograph and found to contain .1231 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is October 17, 2019 The expiration date for this lot number is October 16, 2021 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Alma Palmer, Operations Manager
RepCo Marketing Co.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II

GARRETT HOLMAN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service, and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 506.111 through 506.119 RSMo.

DATE 3/9/2020

NUMBER 200126

EXPIRES 3/9/2022

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LABORATORY

MO-5909774-05105



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator: **HOLMAN, GARRETT**
Permit No: **200126**
Date Issued **3/9/2020** Date Expires **3/9/2022**

