

**RECEIVED**

By Tracy Crews at 7:59 am, Jul 12, 2021

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 087959	NAME OF AGENCY ST CHARLES COUNTY PD	DATE OF INSPECTION 07/09/2021
LOCATION OF INSTRUMENT (STREET AND CITY) 101 SHERIFF DIERKER CT., OFALLON, MO 63366		TIME OF INSPECTION 8:37 am

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument. DIGITAL READOUT (ALL ELEMENTS OPERATIONAL) TEMPERATURE OF ALCO SENSOR (10°C - 40°C) PRINTER WORKING PROPERLY TIME AND DATE DISPLAYING PROPERLY**BREATH ALCOHOL ACCURACY STANDARDS** SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE STANDARD SUPPLIER INTOXIMETERS LOT # AG109701 EXP. DATE 04/07/2023 SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIM. SN SIM. NIST EXP DATE CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVETEST 1 .078TEST 2 .077TEST 3 .078 RFI DETECTOR OPERATING**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	(0-.04)	(.05-.09) <u>1</u>	(.10-.14)	(.15-.19)	(OVER .19)
----------	---------	--------------------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Printer (serial# 096.3580.953) was mailed to Missouri Safety Center where the printer battery was replaced by Matthew Bond.

**INSPECTING OFFICER**

SIGNATURE

Col. Johnson 663

PRINT NAME

MIKE JOHNSON

TYPE II PERMIT NUMBER/EXPIRATION DATE

200198 / 07-08-2022

TELEPHONE NUMBER

(636) 949-3000

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 087959  
Version no: 532B

TEST RECORD 00269

Temp Date Time 210L <sup>s/</sup>

Air Blank:  
07/09/21 08:37 .000  
Calibration Check:  
25 07/09/21 08:37 .078

Subject Name

Test #1

Subject I.D.

Operator Name, I.D.

Cpl. Johnson 200148

Location

SCCPD

AS IV Serial no: 087959  
Version no: 532B

TEST RECORD 00270

Temp Date Time 210L <sup>s/</sup>

Air Blank:  
07/09/21 08:41 .000  
Calibration Check:  
25 07/09/21 08:41 .077

Subject Name

Test #2

Subject I.D.

Operator Name, I.D.

Cpl. Johnson 200148

Location

SCCPD

AS IV Serial no: 087959  
Version no: 532B

TEST RECORD 00271

Temp Date Time 210L <sup>s/</sup>

Air Blank:  
07/09/21 08:43 .000  
Calibration Check:  
26 07/09/21 08:43 .078

Subject Name

Test #3

Subject I.D.

Operator Name, I.D.

Cpl. Johnson 200148

Location

SCCPD

AS IV Serial no: 087959  
Version no: 532B

TEST RECORD 00272

Temp Date Time 210L <sup>s/</sup>

VOID: RFI  
12 07/09/21 08:44

Subject Name

RFI Test

Subject I.D.

Operator Name, I.D.

Cpl. Johnson 200148

Location

SCCPD

AS IV Serial no: 087959  
Version no: 532B

TEST RECORD 00273

Temp Date Time 210L <sup>s/</sup>

Air Blank:  
07/09/21 08:45 .000  
Subject Test: Auto  
27 07/09/21 08:45 .000

Subject Name

Self-Test

Subject I.D.

Operator Name, I.D.

Cpl. Johnson 200148

Location

SCCPD



Airgas USA LLC (LAB)  
3500 Bernard Street  
St. Louis, Mo. 63103  
Ph: (314) 533-3100  
Fax: (314) 533-7328

### Certificate of Analysis

Customer Name  
*Exclusive Supplier*  
Intoximeters, Inc.  
2081 Craig Road  
St. Louis, Mo 63146

Test Date: 7-Apr-2021

Lot # AG109701 Model 108cacd

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
7-Apr-2023	108	Ethanol Nitrogen	0.082 ± 0.002 BrAC (223 ppm) Balance


Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

Analytical Method: NDIR

Digitally signed by Quality Control  
Date: 2021.04.07 17:42:01 -05:00  
Reason: Dry gas standard certification of analysis  
Location: Airgas USA LLC (Lab)

Approved for Release:   
Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06  
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM



**PERMIT  
 TYPE II**

**MIKE JOHNSON**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/8/2020

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 200198

EXPIRES 7/8/2022

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator JOHNSON, MIKE  
 Permit No 200198  
 Date Issued 7/8/2020 Date Expires 7/8/2022