



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 087960	NAME OF AGENCY Marceline Police Department	DATE OF INSPECTION 10/29/2021
-----------------------------	---	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 123 E. Santa Fe., Marceline, Mo, 64658	TIME OF INSPECTION 9:24 pm
--	-------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Intoximeters Inc. LOT # AG104202 EXP. DATE 10/11/2022

SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIM. SN SIM. NIST EXP DATE

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ➡ .097	TEST 2 ➡ .098	TEST 3 ➡ .097
---------------	---------------	---------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14) 1	(.15-.19)	(OVER .19)
----------	---------	-----------	-------------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE 	PRINT NAME William Barger
TYPE II PERMIT NUMBER/EXPIRATION DATE 210056/04/06/2023	TELEPHONE NUMBER (660) 376-3556

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

MARCELINE POLICE DEPARTMENT-BREATH ALCOHOL PROGRAM

AS IV Serial no: 087960
Version no: 532B

TEST RECORD 00389 s/
Temp Date Time 210L

Air Blank:
10/29/21 21:24 .000
Calibration Check:
24 10/29/21 21:24 .097

Subject Name
TEST # 1
Subject I.D.

Operator Name, I.D.
W. BARGER #210056
Location
123 E. SANTA FE
MARCELINE, MO
64658

AS IV Serial no: 087960
Version no: 532B

TEST RECORD 00390 s/
Temp Date Time 210L

Air Blank:
10/29/21 21:26 .000
Calibration Check:
24 10/29/21 21:26 .098

Subject Name
TEST # 2
Subject I.D.

Operator Name, I.D.
W. BARGER #210056
Location
123 E. SANTA FE
MARCELINE, MO
64658

AS IV Serial no: 087960
Version no: 532B

TEST RECORD 00391 s/
Temp Date Time 210L

Air Blank:
10/29/21 21:28 .000
Calibration Check:
24 10/29/21 21:28 .097

Subject Name
TEST # 3
Subject I.D.

Operator Name, I.D.
W. BARGER #210056
Location
123 E. SANTA FE
MARCELINE, MO
64658

AS IV Serial no: 087960
Version no: 532B

TEST RECORD 00392 s/
Temp Date Time 210L

VOID: RFI
12 10/29/21 21:30

Subject Name
RFI TEST
Subject I.D.

Operator Name, I.D.
W. BARGER #210056
Location
123 E. SANTA FE
MARCELINE, MO
64658

TEST# 1

TEST #2

TEST #3

RFI TEST

AS-IV S/N: 087960 Type II Permit Holder: William Barger #210056-Exp: 04/06/2023 Date of Maintenance Report: 10/29/2021

Marceline Police Department AS-IV Monthly Maintenance Report



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 11-Feb-2021

Lot # AG104202 Model 34cacc

Exp. Date	Cyl. Type	Component	Certified Concentration
11-Oct-2022	34	Ethanol Nitrogen	0.100 ± 2% BrAC (272 ppm) Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm
<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
 Date: 2021.02.11 17:38:27 -06:00
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)

Approved for Release:

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
WILLIAM BARGER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/6/2021

NUMBER 210056

EXPIRES 4/6/2023

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

 **STATE OF MISSOURI**
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator BARGER, WILLIAM
 Permit No 210056
 Date Issued 4/6/2021 Date Expires 4/6/2023

