

AS IV Serial no: 087964
Version no: 532B

TEST RECORD 00694
Temp Date Time 210L
s/

Air Blank: 08/22/21 05:04 .000
Calibration Check: 22 08/22/21 05:04 .079

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 087964
Version no: 532B

TEST RECORD 00695
Temp Date Time 210L
s/

Air Blank: 08/22/21 05:06 .000
Calibration Check: 22 08/22/21 05:06 .078

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 087964
Version no: 532B

TEST RECORD 00696
Temp Date Time 210L
s/

Air Blank: 08/22/21 05:08 .000
Calibration Check: 23 08/22/21 05:08 .078

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 087964
Version no: 532B

TEST RECORD 00697
Temp Date Time 210L
s/

VOID: RFI 12 08/22/21 05:09

Subject Name

Subject I.D.

Operator Name, I.D.

Location



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 2-Oct-2018

Lot # AG827002 Model 108cacc

<u>Exp. Date</u> 27-Sep-2020	<u>Cyl. Type</u> 108	<u>Component</u> Ethanol Nitrogen	<u>Certified Concentration</u> 0.080 ± 0.002 BrAC (208 ppm) Balance
---------------------------------	-------------------------	---	---

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

<u>Serial No.</u>	<u>Concentration</u>	<u>Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
 Date: 2018.10.03 10:02:04 -05:00
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)

Approved for Release: _____

Rod Marsala

 Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
ALEXANDER K. STIMSON

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/6/2021

NUMBER 210092

EXPIRES 5/6/2023

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)

 STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator STIMSON, ALEXANDER
 Permit No 210092
 Date Issued 5/6/2021 Date Expires 5/6/2023

