



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY

**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 097411	PRINTER SN 03A.2436.096	DATE OF INSPECTION 11/02/2021
LOCATION OF INSTRUMENT (STREET AND CITY) 4001 NE Lakewood Way Ct., Lee's Summit, Mo 64064		TIME OF INSPECTION 9:43 pm

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C) 22°
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

<input type="checkbox"/> SIMULATOR SOLUTION	<input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
STANDARD SUPPLIER <u>Intoximeters</u> LOT # <u>AG006306</u> EXP. DATE <u>03-03-2022</u>	
SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIMULATOR SN _____ SIMULATOR EXP DATE _____	

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .101	TEST 2 ← .101	TEST 3 ← .101
---------------	---------------	---------------

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	4	(0-.04)	0	(.05-.09)	1	(.10-.14)	3	(.15-.19)	1	(OVER .19)	2
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).  
Instrument meets all DHSS standards and guideline.

<b>INSPECTING OFFICER</b>	
SIGNATURE 	PRINT NAME Sean Plain
TYPE II PERMIT NUMBER/EXPIRATION DATE 200054 01/10/2022	TELEPHONE NUMBER (816) 795-1960

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
2875 James Boulevard  
Poplar Bluff, MO 63901

68 70 Serial # 1 09744  
1941 01100 1940

68 70 Serial # 1 09744  
1941 01100 1940

68 70 Serial # 1 09744  
1941 01100 1940

Monthly Manual

Dep. S. Plain #107

JCSO GHO

68 70 Serial # 1 09744  
1941 01100 1940

68 70 Serial # 1 09744  
1941 01100 1940

68 70 Serial # 1 09744  
1941 01100 1940

Monthly Manual

Dep. S. Plain #107

JCSO GHO

68 70 Serial # 1 09744  
1941 01100 1940

68 70 Serial # 1 09744  
1941 01100 1940

68 70 Serial # 1 09744  
1941 01100 1940

Monthly Manual

Dep. S. Plain #107

JCSO GHO

68 70 Serial # 1 09744  
1941 01100 1940

68 70 Serial # 1 09744  
1941 01100 1940

68 70 Serial # 1 09744  
1941 01100 1940

Monthly Manual

Dep. S. Plain #107

JCSO GHO





STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**SEAN PLAIN**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 1/10/2020

NUMBER 200054

EXPIRES 1/10/2022

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



**STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator PLAIN, SEAN  
 Permit No 200054  
 Date Issued 1/10/2020 Date Expires 1/10/2022

