

AS IV Serial no: 097423
Version no: 532B

TEST RECORD 00287

Temp Date Time ^{s/} 210L

Air Blank:
11/01/21 13:37 .000
Calibration Check:
21 11/01/21 13:37 .103

Subject Name

Maint Test #1
Subject I.D.

Operator Name, I.D.

Duffie #1325
Location

Troop C HQ

AS IV Serial no: 097423
Version no: 532B

TEST RECORD 00288

Temp Date Time ^{s/} 210L

Air Blank:
11/01/21 13:39 .000
Calibration Check:
22 11/01/21 13:39 .102

Subject Name

Maint Test #2
Subject I.D.

Operator Name, I.D.

Duffie #1325
Location

Troop C HQ

AS IV Serial no: 097423
Version no: 532B

TEST RECORD 00289

Temp Date Time ^{s/} 210L

Air Blank:
11/01/21 13:41 .000
Calibration Check:
23 11/01/21 13:41 .100

Subject Name

Maint Test #3
Subject I.D.

Operator Name, I.D.

Duffie #1325
Location

Troop C HQ

AS IV Serial no: 097423
Version no: 532B

TEST RECORD 00290

Temp Date Time ^{s/} 210L

VOID: NFI
12 11/01/21 13:42

Subject Name

RET Test
Subject I.D.

Operator Name, I.D.

Duffie #1325
Location

Troop C HQ

RepCo Marketing Co
3101-188 Stony Brook Drive
Raleigh, NC 27604
919-876-5480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing Co.

LOT NUMBER: 20001

EXPIRATION DATE: October 7, 2022 at 11:59 p.m.

RepCo Marketing Co. certifies the following:


RepCo Marketing Co. prepared, tested and supplied Lot Number 20001 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by ISO 17025 and ISO 17034 accredited institution Alcohol Countermeasure Systems, using NIST standards. Random samples were analyzed by Alcohol Countermeasure Systems utilizing a gas chromatograph and found to contain .1227 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is October 8, 2020 The expiration date for this lot number is October 7, 2022 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Alma Palmer, Operations Manager
RepCo Marketing Co.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
DUSTIN J DUFFIE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/4/2020

NUMBER 200164

EXPIRES 5/4/2022

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator DUFFIE, DUSTIN
Permit No 200164
Date Issued 5/4/2020 **Date Expires** 5/4/2022

