

AS IV Serial no: 099363
Version no: 532B

TEST RECORD 00703

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
03/07/21 09:55 .000
Calibration Check:
11 03/07/21 09:55 .104

Subject Name

MAINT TEST

Subject I.D.

Operator Name, I.D.

S.D. Fletcher 661

Location

AS IV Serial no: 099363
Version no: 532B

TEST RECORD 00704

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
03/07/21 09:58 .000
Calibration Check:
14 03/07/21 09:58 .103

Subject Name

MAINT TEST

Subject I.D.

Operator Name, I.D.

S.D. Fletcher 661

Location

AS IV Serial no: 099363
Version no: 532B

TEST RECORD 00705

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
03/07/21 10:04 .000
Calibration Check:
16 03/07/21 10:04 .104

Subject Name

MAINT TEST

Subject I.D.

Operator Name, I.D.

S.D. Fletcher 661

Location

AS IV Serial no: 099363
Version no: 532B

TEST RECORD 00706

Temp	Date	Time	s/ 210L
------	------	------	------------

VOID: RFI
12 03/07/21 10:06

Subject Name

RFI CHECK

Subject I.D.

Operator Name, I.D.

S.D. Fletcher 661

Location

CERTIFICATE OF ANALYSIS**MANUFACTURER AND SUPPLIER: RepCo Marketing Co.****LOT NUMBER: 19002****EXPIRATION DATE: October 16, 2021 at 11:59 p.m.**

RepCo Marketing Co. certifies the following:


RepCo Marketing Co. prepared, tested and supplied Lot Number 19002 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by ISO 17025 and ISO 17034 accredited institution Alcohol Countermeasure Systems, using NIST standards. Random samples were analyzed by Alcohol Countermeasure Systems utilizing a gas chromatograph and found to contain .1231 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is October 17, 2019 The expiration date for this lot number is October 16, 2021 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.


Alma Palmer, Operations Manager
RepCo Marketing Co.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



**PERMIT
TYPE II**

SHANNON D BLEDSOE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/16/2019

NUMBER 290289

EXPIRES 12/16/2021

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

 **STATE OF MISSOURI**
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator BLEDSOE, SHANNON
Permit No 290289
Date Issued 12/16/2019 **Date Expires** 12/16/2021

