



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
By Tracy Crews at 2:17 pm, May 17, 2021

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 100284	NAME OF AGENCY Columbia Police Department	DATE OF INSPECTION 05/01/2021
LOCATION OF INSTRUMENT (STREET AND CITY) 600 East Walnut Columbia		TIME OF INSPECTION 7:06 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

<input type="checkbox"/> SIMULATOR SOLUTION	<input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>Intoximeters</u>	LOT # <u>AG020401</u> EXP. DATE <u>07/22/2022</u>
<input type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C)	SIM. SN _____ SIM. NIST EXP DATE _____

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .098	TEST 2 ← .098	TEST 3 ← .098
----------------------	----------------------	----------------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 1	(0-.04) 0	(.05-.09) 1	(.10-.14) 0	(.15-.19) 0	(OVER .19) 1
-------------------	------------------	--------------------	--------------------	--------------------	---------------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Monthly Maintenance

INSPECTING OFFICER	
SIGNATURE <i>Mark D. Hoehne</i>	PRINT NAME Mark D. Hoehne
TYPE II PERMIT NUMBER/EXPIRATION DATE 200187 06/15/2022	TELEPHONE NUMBER (573-874-7585)

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 100284
Version no: 532B

TEST RECORD - REPRINT

TEST RECORD 00273

Temp Date Time ^{s/} 210L

Air Blank:
05/01/21 19:07 .000
Calibration Check:
24 05/01/21 19:07 .000

Subject Name

Test # 1

Subject I.D.

Hochne, Mark D. 2078
Operator Name, I.D.

Location

AS IV Serial no: 100284
Version no: 532B

TEST RECORD 00274

Temp Date Time ^{s/} 210L

Air Blank:
05/01/21 19:11 .000
Calibration Check:
24 05/01/21 19:11 .000

Subject Name

Test # 2

Subject I.D.

Hochne, Mark D. 2078
Operator Name, I.D.

Location

AS IV Serial no: 100284
Version no: 532B

TEST RECORD 00275

Temp Date Time ^{s/} 210L

Air Blank:
05/01/21 19:14 .000
Calibration Check:
24 05/01/21 19:14 .000

Subject Name

Test # 3

Subject I.D.

Hochne, Mark D. 2078
Operator Name, I.D.

Location

AS IV Serial no: 100284
Version no: 532B

TEST RECORD 00276

Temp Date Time ^{s/} 210L

VOID: RFI
12 05/01/21 19:15

Subject Name

RFI Check

Subject I.D.

Hochne, Mark D. 2078
Operator Name, I.D.

Location

AS IV Serial no: 100284
Version no: 532B

TEST RECORD 00277

Temp Date Time ^{s/} 210L

Air Blank:
05/01/21 19:17 .000
Subject Test: Auto
23 05/01/21 19:17 .000

Subject Name

Self Test

Subject I.D.

Hochne, Mark D. 2078
Operator Name, I.D.

Location

Customer Name
Exclusive Supplier
InfoAnalysts, Inc
2931 Craig Road
St Louis Mo 63145

Lot # AG020401 Model 108caed

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
22-Jul-2022	108	Ethanol Nitrogen	0.100 ± 2% BrAC (272 ppm) Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010695	208.3 ppm
EB0010561	103.6 ppm	EB0010662	104.2 ppm
EB0010881	52.12 ppm	EB0010579	52.81 ppm

<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC434668	800.0 ppm	0056649	390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
Date: 2020.07.22 08:29:45 -05:00
Reason: Dry gas standard certification of analysis
Location: Algas USA LLC (Lsh)

Approved for Release:


Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07

with the following information and signature of operator, date, instrument number, calibration date and field number of instrument, and the following information:

ALCO-SENSOR IV WITH PRINTER, INTOX DMUE

for determination of the alcohol content of blood from a sample of expired air. Permit issued under the provisions of sections 320 through 377.041, RSMo and 306.111 through 306.119 RSMo

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
INSTRUMENT OPERATOR PROGRAM

INSTRUMENT OPERATOR CARD

The instrument operator must sign and date this card at the time of the determination of the alcohol content in breath from a person in Missouri.

Operator: **HOUSTON, MARK**

Permit No: **200187**

Date Issued: **5/15/2020** Date Expires: **12/15/2022**

