

AS IV Serial no: 100286
Version no: 532B

TEST RECORD 00921
Temp Date Time 210L
a/

Air Blank:
01/04/21 03:07 .000
Calibration Check:
20 01/04/21 03:07 .103

Subject Name
Blank

Subject I.D.
Test 1

Operator Name, I.D.
Centurion / 200163

Location
602 S. Franklin

Cuba

AS IV Serial no: 100286
Version no: 532B

TEST RECORD 00922
Temp Date Time 210L
a/

Air Blank:
01/04/21 03:08 .000
Calibration Check:
21 01/04/21 03:08 .102

Subject Name
Blank

Subject I.D.
Test 2

Operator Name, I.D.
Centurion / 200163

Location
602 S. Franklin

Cuba

AS IV Serial no: 100286
Version no: 532B

TEST RECORD 00923
Temp Date Time 210L
a/

Air Blank:
01/04/21 03:11 .000
Calibration Check:
21 01/04/21 03:11 .102

Subject Name
Blank

Subject I.D.
Test 3

Operator Name, I.D.
Centurion / 200163

Location
602 S. Franklin

Cuba

AS IV Serial no: 100286
Version no: 532B

TEST RECORD 00924
Temp Date Time 210L
a/

VOID: RFI
12 01/04/21 03:12

Subject Name
Blank

Subject I.D.
RFI

Operator Name, I.D.
Centurion / 200163

Location
602 S. Franklin

Cuba



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **20190** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **April 8, 2020**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1206%** (w/vol) ethyl alcohol. The expiration date for this lot number is **April 6, 2022** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN02271802 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II

MICHAEL S CENTUNZI

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/4/2020

NUMBER 200163

EXPIRES 5/4/2022

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (5-10)

LAB-4 (R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator CENTUNZI, MICHAEL
Permit No 200163
Date Issued 5/4/2020 Date Expires 5/4/2022



