



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 100288	NAME OF AGENCY St. Louis County Police Department	DATE OF INSPECTION 07/09/2021
LOCATION OF INSTRUMENT (STREET AND CITY) 16860 Main St (6th PCT)		TIME OF INSPECTION 8:44 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Intoximeters LOT # AG015503 EXP. DATE 06/03/2022

SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIM. SN SIM. NIST EXP DATE

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .099	TEST 2 .099	TEST 3 .098
--------------	--------------	--------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE 	PRINT NAME PO NEUMAN, DSN 3072
TYPE II PERMIT NUMBER/EXPIRATION DATE 210074 04/06/2023	TELEPHONE NUMBER (636) 529-8210

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 100288
Version no: 532B

TEST RECORD 00618

Temp Date Time ^{a/} 210L

Air Blank:
07/09/21 08:44 .000
Calibration Check:
21 07/09/21 08:44 .099

Subject Name

Test #1

Subject I.D.

N/A

Operator Name, I.D.

PO D. Newman #3072

Location

116860 Main St

(6th PCT)

AS IV Serial no: 100288
Version no: 532B

TEST RECORD 00619

Temp Date Time ^{a/} 210L

Air Blank:
07/09/21 08:47 .000
Calibration Check:
22 07/09/21 08:47 .099

Subject Name

Test #2

Subject I.D.

N/A

Operator Name, I.D.

PO D. Newman #3072

Location

116860 Main St

(6th PCT)

AS IV Serial no: 100288
Version no: 532B

TEST RECORD 00620

Temp Date Time ^{a/} 210L

Air Blank:
07/09/21 08:50 .000
Calibration Check:
23 07/09/21 08:50 .098

Subject Name

Test #3

Subject I.D.

N/A

Operator Name, I.D.

PO D. Newman #3072

Location

116860 Main St

(6th PCT)

AS IV Serial no: 100288
Version no: 532B

TEST RECORD 00621

Temp Date Time ^{a/} 210L

VOID: RFI
12 07/09/21 08:52

Subject Name

RFI!

Subject I.D.

N/A

Operator Name, I.D.

PO D. Newman #3072

Location

116860 Main St

(6th PCT)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

**PERMIT
 TYPE II**

DAWN M. NEUMAN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/6/2021

NUMBER 210074

EXPIRES 4/6/2023

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES