



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 102454	PRINTER SN 095.3583.044	DATE OF INSPECTION 06/18/2021
LOCATION OF INSTRUMENT (STREET AND CITY) Monett Police Department, 1901 E. Cleveland Ave, Monett, MO, 65708		TIME OF INSPECTION 5:06 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER Intoximeters LOT # AG015503 EXP. DATE 06/22/2022
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIMULATOR SN _____ SIMULATOR EXP DATE _____

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 \blacktriangleleft .099	TEST 2 \blacktriangleleft .100	TEST 3 \blacktriangleleft 0.100
----------------------------------	----------------------------------	-----------------------------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	2	(.15-.19)	0	(OVER .19)	1
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER	
SIGNATURE 	PRINT NAME Trent Gold
TYPE II PERMIT NUMBER/EXPIRATION DATE 210049 3-30-23	TELEPHONE NUMBER 417-235-4241

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901

AS IV Serial no: 102454
Version no: 532B

TEST RECORD 00523

Temp Date Time 210L

Air Blank:
06/18/21 05:07 .000
Calibration Check:
19 06/18/21 05:07 .099

Subject Name
Test 1
Subject I.D.

Operator Name, I.D. ³⁻³⁰⁻²³
Gold, Trent 210049
Location
Morett PD

1901 E Cleveland Ave
Morett MO 65708

AS IV Serial no: 102454
Version no: 532B

TEST RECORD 00524

Temp Date Time 210L

Air Blank:
06/18/21 05:08 .000
Calibration Check:
19 06/18/21 05:08 .100

Subject Name
Test 2
Subject I.D.

Operator Name, I.D. ³⁻³⁰⁻²³
Gold, Trent 210049
Location
Morett PD

AS IV Serial no: 102454
Version no: 532B

TEST RECORD 00525

Temp Date Time 210L

Air Blank:
06/18/21 05:09 .000
Calibration Check:
20 06/18/21 05:09 .100

Subject Name
Test 3
Subject I.D.

Operator Name, I.D. ³⁻³⁰⁻²³
Gold, Trent 210049
Location
Morett PD

AS IV Serial no: 102454
Version no: 532B

TEST RECORD 00526

Temp Date Time 210L

VOID: RFI
12 06/18/21 05:10

Subject Name
RFI
Subject I.D.

Operator Name, I.D. ³⁻³⁰⁻²³
Gold, Trent 210049
Location
Morett PD

AS IV Serial no: 102454
Version no: 532B

TEST RECORD 00527

Temp Date Time 210L

Air Blank:
06/18/21 05:12 .000
Subject Test: Auto
22 06/18/21 05:12 .000

Subject Name
Operational tes
Subject I.D.

Operator Name, I.D. ³⁻³⁰⁻²³
Gold, Trent 210049
Location
Morett PD



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

Certificate of Analysis

Customer Name

Exclusive Supplier

Intoximeters, Inc.

2081 Craig Road

St. Louis, Mo 63146

Test Date: 8-Jun-2020

Lot # AG015503 Model 108cacd

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
3-Jun-2022	108	Ethanol	0.100 ± 2% BrAC (272 ppm)
		Nitrogen	Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm
<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC434668	800.0 ppm	0056649	390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
Date: 2020.06.10 14:11:39 -05:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Approved for Release:



Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
TRENT GOLD

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.


DATE 3/30/2021



DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 210049

EXPIRES 3/30/2023



DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (5-10)

LAB-4 (R5-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator **GOLD, TRENT**
Permit No **210049**
Date Issued **3/30/2021** Date Expires **3/30/2023**

