



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
 By Tracy Crews at 9:43 am, Jan 05, 2022

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 102469	PRINTER SN 095.353.193	DATE OF INSPECTION 01/04/2022
LOCATION OF INSTRUMENT (STREET AND CITY) Lafayette County Sheriff 107 S. 11th Lexington, Mo. 64067		TIME OF INSPECTION 3:41 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Labs LOT # 20190 EXP. DATE 04/06/2022

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN SD2275 SIMULATOR EXP DATE 01/11/2022

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .098

TEST 2 .097

TEST 3 .099

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Meets DOH Standards

INSPECTING OFFICER

SIGNATURE

PRINT NAME
Dale L. Cox

TYPE II PERMIT NUMBER/EXPIRATION DATE
200261 10/02/2022

TELEPHONE NUMBER
(660) 259-3622

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901

55 IN Serial no: 102469
 Version no: 532B
 TEST RECORD 00743
 Temp Date Time 210L
 Air Blank: 01/04/22 15:43 .000
 Calibration Check: 24 01/04/22 15:43 .097
 Subject Name
 Col-Check
 Subject I.D.

Operator Name- I.D.
 COK 200261
 Location
 LLSO

55 IN Serial no: 102469
 Version no: 532B
 TEST RECORD 00744
 Temp Date Time 210L
 Air Blank: 01/04/22 15:45 .000
 Calibration Check: 24 01/04/22 15:45 .099
 Subject Name
 Col-Check
 Subject I.D.

Operator Name- I.D.
 COK 200261
 Location
 LLSO

55 IN Serial no: 102469
 Version no: 532B
 TEST RECORD 00745
 Temp Date Time 210L
 Air Blank: 01/04/22 15:45 .000
 Calibration Check: 24 01/04/22 15:45 .099
 Subject Name
 Col-Check
 Subject I.D.

Operator Name- I.D.
 COK 200261
 Location
 LLSO

55 IN Serial no: 102469
 Version no: 532B
 TEST RECORD 00742
 Temp Date Time 210L
 Air Blank: 01/04/22 15:41 .000
 Calibration Check: 22 01/04/22 15:41 .000
 Subject Name
 Ad-Check
 Subject I.D.

Operator Name- I.D.
 COK 200261
 Location
 LLSO



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **20190** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **April 8, 2020**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1206%** (w/vol) ethyl alcohol. The expiration date for this lot number is **April 6, 2022** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN02271802 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
DALE L COX

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 10/2/2020

NUMBER 200261

EXPIRES 10/2/2022

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4.(R6-10)

MO 580:0771 (6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator COX, DALE
Permit No 200261
Date Issued 10/2/2020 Date Expires 10/2/2022

