



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 104641	NAME OF AGENCY O'Fallon PD	DATE OF INSPECTION 6-18-2021
LOCATION OF INSTRUMENT (STREET AND CITY) 1019 Bryan Road, O'Fallon, MO		TIME OF INSPECTION 0128

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

<input type="checkbox"/> SIMULATOR SOLUTION	<input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>Intoximeters</u>	LOT # <u>AG919902</u> EXP. DATE <u>07-18-2021</u>
<input type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____	SIM. SN _____ SIM. NIST EXP DATE _____

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .079	TEST 2 ← .078	TEST 3 ← .078
----------------------	----------------------	----------------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0	(0-.04) 0	(.05-.09) 0	(.10-.14) 0	(.15-.19) 0	(OVER .19) 0
-------------------	------------------	--------------------	--------------------	--------------------	---------------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

N/A

INSPECTING OFFICER	
SIGNATURE 	PRINT NAME Chris Sinnokrak
TYPE II PERMIT NUMBER/EXPIRATION DATE 200307 / 12-21-2022	TELEPHONE NUMBER (636)240-3200

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 104641
Version no: 532B

TEST RECORD 01098

Temp Date Time 210L ^{s/}

Air Blank:
06/18/21 01:31 .000
Calibration Check:
21 06/18/21 01:31 .079

Subject Name

TEST 1

Subject I.D.

Operator Name, I.D.

SINNOORAK 334

Location

AS IV Serial no: 104641
Version no: 532B

TEST RECORD 01099

Temp Date Time 210L ^{s/}

Air Blank:
06/18/21 01:32 .000
Calibration Check:
22 06/18/21 01:32 .078

Subject Name

TEST 2

Subject I.D.

Operator Name, I.D.

SINNOORAK 334

Location

AS IV Serial no: 104641
Version no: 532B

TEST RECORD - REPRINT

TEST RECORD 01100

Temp Date Time 210L ^{s/}

Air Blank:
06/18/21 01:34 .000
Calibration Check:
23 06/18/21 01:34 .078

Subject Name

TEST 3

Subject I.D.

Operator Name, I.D.

SINNOORAK 334

Location

AS IV Serial no: 104641
Version no: 532B

TEST RECORD 01101

Temp Date Time 210L ^{s/}

VOID: RFI
12 06/18/21 01:36

Subject Name

RFI

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 104641
Version no: 532B

TEST RECORD 01102

Temp Date Time 210L ^{s/}

Air Blank:
06/18/21 01:38 .000
Subject Test: Auto
25 06/18/21 01:38 .000

Subject Name

SOBER TEST

Subject I.D.

Operator Name, I.D.

Location



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

Certificate of Analysis

Customer Name

Exclusive Supplier

Intoximeters, Inc.

2081 Craig Road

St. Louis, Mo 63146

Test Date: 22-Jul-2019

Lot # AG919902 **Model** 108caccd

Exp. Date

18-Jul-2021

Cyl. Type

108

Component

Ethanol

Nitrogen

Certified Concentration

0.080 ± 0.002 BrAC (218 ppm)

Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.

EB0010581

Concentration

392.1 ppm

EB0010570

259.8 ppm

EB0010285

208.0 ppm

EB0010561

103.6 ppm

EB0010681

52.12 ppm

RGM Serial No.

EB0010603

Concentration

393.0 ppm

EB0010559

258.2 ppm

EB0010595

208.3 ppm

EB0010562

104.2 ppm

EB0010579

52.81 ppm

CRM Serial No.

CC434668

Concentration

800.0 ppm

CC234503

253.0 ppm

CRM Serial No.

0056649

Concentration

390.1 ppm

0056662

150.2 ppm

Analytical Method:

NDIR

Digitally signed by Quality Control
Date: 2019.07.22 13:17:50 -05:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Approved for Release:

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06

ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
CHRIS SINNOKRAK

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/21/2020

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 200307

EXPIRES 12/21/2022

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator SINNOKRAK, CHRIS
Permit No 200307
Date Issued 12/21/2020 **Date Expires** 12/21/2022