

RECEIVED

By Tracy Crews at 12:18 pm, Aug 16, 2021

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 104658 PRINTER SN 096.3580.740 DATE OF INSPECTION 08/14/2021

LOCATION OF INSTRUMENT (STREET AND CITY) 513 East Grand Avenue, Cameron, MO, 64429 TIME OF INSPECTION 7:45 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument. DIGITAL READOUT (ALL ELEMENTS OPERATIONAL) TEMPERATURE OF ALCO SENSOR (10°C - 40°C) PRINTER WORKING PROPERLY TIME AND DATE DISPLAYING PROPERLY**BREATH ALCOHOL ACCURACY STANDARDS** SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE STANDARD SUPPLIER Repco LOT # 20001 EXP. DATE 10/07/2022 SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN MP2488 SIMULATOR EXP DATE 02/09/2022 CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVETEST 1 .100TEST 2 .099TEST 3 .098 RFI DETECTOR OPERATINGINDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 1 (0-.04) 0 (.05-.09) 0 (.10-.14) 1 (.15-.19) 2 (OVER .19) 0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE

PRINT NAME

Corporal M. J. Miller #1418

TYPE II PERMIT NUMBER/EXPIRATION DATE

200292

12/08/2022

TELEPHONE NUMBER

(816) 387-2345

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AS TV Serial no: 18480
Version no: 3228

WEST RECORD 80279

Temp Date Time 2:00

Mr Blank
08/14/71 07:55 .00
Calibration Check
22 45/16/71 07:54 .00

Subject Name

MAINTENANCE #1

N/A

Operator Name: J.L.

M.J. MILLER #1418

513 E GRAND AVE

CAMERON, MO, 64429

M.J. Miller

AS TV Serial no: 18480
Version no: 3228

WEST RECORD 80279

Temp Date Time 2:00

Mr Blank
08/14/71 27:54 .00
Calibration Check
22 45/16/71 07:54 .00

Subject Name

MAINTENANCE #2

N/A

Operator Name: J.L.

M.J. MILLER #1418

513 E GRAND AVE

CAMERON, MO, 64429

M.J. Miller

AS TV Serial no: 18480
Version no: 3228

WEST RECORD 80291

Temp Date Time 2:00

Mr Blank
08/14/71 07:53 .00

Subject Name

MAINTENANCE REF

N/A

Operator Name: J.L.

M.J. MILLER #1418

513 E GRAND AVE

CAMERON, MO, 64429

M.J. Miller

AS TV Serial no: 18480
Version no: 3228

WEST RECORD 80292

Temp Date Time 2:00

Mr Blank
08/14/71 07:54 .00
Calibration Check
22 45/16/71 07:54 .00

Subject Name

MAINTENANCE #3

N/A

Operator Name: J.L.

M.J. MILLER #1418

513 E GRAND AVE

CAMERON, MO, 64429

M.J. Miller

RepCo Marketing Co
3101-188 Story Brook Drive
Raleigh, NC 27604
919-876-5480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing Co.

LOT NUMBER: 20001

EXPIRATION DATE: October 7, 2022 at 11:59 p.m.

RepCo Marketing Co. certifies the following:


RepCo Marketing Co. prepared, tested and supplied Lot Number 20001 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by ISO 17025 and ISO 17034 accredited institution Alcohol Countermeasure Systems, using NIST standards. Random samples were analyzed by Alcohol Countermeasure Systems utilizing a gas chromatograph and found to contain .1227 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is October 8, 2020. The expiration date for this lot number is October 7, 2022 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Alma Palmer, Operations Manager
RepCo Marketing Co.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
MICHAEL J. MILLER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/8/2020

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 200292

EXPIRES 12/8/2022

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580 0771 (6-10)

LAD-4 (16-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri

Operator MILLER, MICHAEL
Permit No 200292
Date Issued 12/8/2020 **Date Expires** 12/8/2022