

RECEIVED

By Tracy Crews at 8:14 am, May 25, 2021



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 105444	NAME OF AGENCY City of Clever Police Department	DATE OF INSPECTION 05/20/2021
LOCATION OF INSTRUMENT (STREET AND CITY) 210 S Clarke Ave, Clever, Missouri 65631		TIME OF INSPECTION 1730 hours

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Laboratories LOT # 21080 EXP. DATE 03/08/2023

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 °C SIM. SN SD2259 SIM. NIST EXP DATE 05/11/2022

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 0.100

TEST 2 0.099

TEST 3 0.100

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0 (0-.04) 0 (.05-.09) 0 (.10-.14) 0 (.15-.19) 0 (OVER .19) 0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Maintenance Report conducted within DHSS rules and regulations.

INSPECTING OFFICER

SIGNATURE [Signature]

PRINT NAME Dillon Petersen

TYPE II PERMIT NUMBER/EXPIRATION DATE 210108 05/18/2023

TELEPHONE NUMBER (417) 743-5109

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 105444
Version no: 532B

TEST RECORD 00740

Temp Date Time ^{s/} 210L

Air Blank:
05/20/21 18:45 .000
Calibration Check:
25 05/20/21 18:45 .100

Subject Name

Test 1

Subject I.D.

Operator Name, I.D.

D Petersen 210108

Location

210 S Clarke Ave

Clever, MO 65631

AS IV Serial no: 105444
Version no: 532B

TEST RECORD 00741

Temp Date Time ^{s/} 210L

Air Blank:
05/20/21 18:46 .000
Calibration Check:
25 05/20/21 18:46 .099

Subject Name

Test 2

Subject I.D.

Operator Name, I.D.

D Petersen 210108

Location

210 S Clarke Ave

Clever, MO 65631

AS IV Serial no: 105444
Version no: 532B

TEST RECORD 00742

Temp Date Time ^{s/} 210L

Air Blank:
05/20/21 18:48 .000
Calibration Check:
26 05/20/21 18:48 .100

Subject Name

Test 3

Subject I.D.

Operator Name, I.D.

D Petersen 210108

Location

210 S Clarke Ave

Clever, MO 65631

AS IV Serial no: 105444
Version no: 532B

TEST RECORD 00743

Temp Date Time ^{s/} 210L

VOID: RFI
12 05/20/21 18:49

Subject Name

RFI Test

Subject I.D.

Operator Name, I.D.

D Petersen 210108

Location

210 S Clarke Ave

Clever, MO 65631

AS IV Serial no: 105444
Version no: 532B

TEST RECORD 00744

Temp Date Time ^{s/} 210L

Air Blank:
05/20/21 18:51 .000
Calibration Check:
26 05/20/21 18:51 .000

Subject Name

Blank Test

Subject I.D.

Operator Name, I.D.

D Petersen 210108

Location

210 S Clarke Ave

Clever, MO 65631



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **21080** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **March 10, 2021**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1214%** (w/vol) ethyl alcohol. The expiration date for this lot number is **March 8, 2023** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03052002 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
DILLON PETERSEN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/18/2021

NUMBER 210108

EXPIRES 5/18/2023

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

 **STATE OF MISSOURI**
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator PETERSEN, DILLON
 Permit No 210108
 Date Issued 5/18/2021 Date Expires 5/18/2023

