



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 105445	NAME OF AGENCY Platte County Sheriff's Office	DATE OF INSPECTION 09/09/2021
LOCATION OF INSTRUMENT (STREET AND CITY) 415 Third Street, Platte City, Missouri 64079		TIME OF INSPECTION 21:13

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Intoximeters, Inc. LOT # AG928303 EXP. DATE 10/10/2021

SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIM. SN _____ SIM. NIST EXP DATE _____

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .101	TEST 2 ← .100	TEST 3 ← .100
---------------	---------------	---------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04) 1	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19) 1
----------	------------------	-----------	-----------	-----------	---------------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE: <i>William Beeler</i>	PRINT NAME: William Beeler
TYPE II PERMIT NUMBER/EXPIRATION DATE: 200168/05-11-2022	TELEPHONE NUMBER: (816) 858-3521

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 105445
Version no: 532B

TEST RECORD 01034

Temp	Date	Time	s/	210L
------	------	------	----	------

Air Blank:
09/09/21 21:13 .000
Calibration Check:
20 09/09/21 21:13 .101

Subject Name

Test 1

Subject I.D.

Operator Name, I.D.

W. Beeler ²⁰⁰¹⁶⁸
5/11/22

Location

415 Third St

Platte City MO

AS IV Serial no: 105445
Version no: 532B

TEST RECORD 01035

Temp	Date	Time	s/	210L
------	------	------	----	------

Air Blank:
09/09/21 21:15 .000
Calibration Check:
21 09/09/21 21:15 .100

Subject Name

Test 2

Subject I.D.

Operator Name, I.D.

W. Beeler ²⁰⁰¹⁶⁸
5/11/22

Location

415 Third St

Platte City MO

AS IV Serial no: 105445
Version no: 532B

TEST RECORD 01036

Temp	Date	Time	s/	210L
------	------	------	----	------

Air Blank:
09/09/21 21:17 .000
Calibration Check:
21 09/09/21 21:17 .100

Subject Name

Test 3

Subject I.D.

Operator Name, I.D.

W. Beeler ²⁰⁰¹⁶⁸
5/11/22

Location

415 Third St

Platte City MO

AS IV Serial no: 105445
Version no: 532B

TEST RECORD 01037

Temp	Date	Time	s/	210L
------	------	------	----	------

VOID: RFI
12 09/09/21 21:19

Subject Name

RFI

Subject I.D.

Operator Name, I.D.

W. Beeler ²⁰⁰¹⁶⁸
5/11/22

Location

415 Third St

Platte City MO

AS IV Serial no: 105445
Version no: 532B

TEST RECORD 01038

Temp	Date	Time	s/	210L
------	------	------	----	------

Air Blank:
09/09/21 21:20 .000
Subject Test: Auto
22 09/09/21 21:20 .000

Subject Name

Sober Sample

Subject I.D.

Operator Name, I.D.

W. Beeler ²⁰⁰¹⁶⁸
5/11/22

Location

415 Third St

Platte City MO



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

**PERMIT
 TYPE II**

WILLIAM BEELER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/11/2020

NUMBER 200168

EXPIRES 5/11/2022

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator BEELER, WILLIAM
Permit No 200168
Date Issued 5/11/2020 **Date Expires** 5/11/2022



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 11-Oct-2019

Lot # AG928303 Model 108cadd

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
10-Oct-2021	108	Ethanol Nitrogen	0.100 ± 2% BrAC (272 ppm) Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC434668	800.0 ppm	0056649	390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
 Date: 2019.10.11 17:27:37 -05:00
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)

Approved for Release: _____

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07