



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 105446	NAME OF AGENCY LONE JACK PD	DATE OF INSPECTION 04/26/2021
LOCATION OF INSTRUMENT (STREET AND CITY) 401 N BYNUM RD, LONE JACK, MO. 64070		TIME OF INSPECTION 1:54 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER GUTH LABORATORIES LOT # 20420 EXP. DATE 09/22/2022

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIM. SN SD2238 SIM. NIST EXP DATE 05/28/2021

CALIBRATION CHECK -- (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 0.098% TEST 2 0.098% TEST 3 0.096%

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	1
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER	
SIGNATURE <i>Zachary Argetsinger</i> #372	PRINT NAME ZACHARY ARGETSINGER
TYPE II PERMIT NUMBER/EXPIRATION DATE 210055 04/05/2023	TELEPHONE NUMBER (816) 697-2417

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

HS IV Serial no: 105446
Version no: 532B

TEST RECORD - REPRINT

TEST RECORD 00844

Temp Date Time 210L

Air Blank: 04/26/21 13:54 .000

Subject Test: Auto 24-04/26/21 13:54 .000

Subject Name

Subject I.D. Dig Readout

Operator Name: I.D.

Location Argetsinger

401 N. Bynum

Lone Jack mo.

HS IV Serial no: 105446
Version no: 532B

TEST RECORD 00845

Temp Date Time 210L

Air Blank: 04/26/21 14:22 .000

Calibration Check: 25 04/26/21 14:22 .098

Subject Name

Subject I.D. Test #1

Operator Name: I.D.

Location Argetsinger

Lone Jack P.D.

HS IV Serial no: 105446
Version no: 532B

TEST RECORD 00846

Temp Date Time 210L

Air Blank: 04/26/21 14:24 .000

Calibration Check: 25 04/26/21 14:24 .098

Subject Name

Subject I.D. Test #2

Operator Name: I.D.

Location Argetsinger

Lone Jack P.D.

HS IV Serial no: 105446
Version no: 532B

TEST RECORD 00847

Temp Date Time 210L

Air Blank: 04/26/21 14:27 .000

Calibration Check: 26 04/26/21 14:27 .096

Subject Name

Subject I.D. Test #3

Operator Name: I.D.

Location Argetsinger

Lone Jack PD

HS IV Serial no: 105446 Version no: 532B	TEST RECORD 00848	Temp Date Time 210L	VOID: RFI 12 04/26/21 14:29	Subject Name <u>RFI Test</u>	Subject I.D.	Operator Name: I.D.	Location <u>Argetsinger</u>	<u>Lone Jack PD</u>
---	-------------------	---------------------	--------------------------------	---------------------------------	--------------	---------------------	--------------------------------	---------------------



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-664-6470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 20420 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on September 23, 2020, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1208% (w/vol) ethyl alcohol. The expiration date for this lot number is September 22, 2022 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03052002 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

ZACHARY A. ARGETSINGER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/5/2021

NUMBER 210055

EXPIRES 4/5/2023

MO 680-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator ARGETSINGER, ZACHARY
Permit No 210055
Date Issued 4/5/2021 Date Expires 4/5/2023