

AS IV Serial no: 105448
Version no: 532B

TEST RECORD 01474

Temp Date Time 210L

Air Blank:
07/01/21 10:04 .000
Calibration Check:
22 07/01/21 10:04 .098

Subject Name

Test #1

Subject I.D.

N/A

Operator Name, I.D.

M. Nelson #200210

Location

Clinton P.D.

101 E. Ohio

Clinton, MO 64735

AS IV Serial no: 105448
Version no: 532B

TEST RECORD 01475

Temp Date Time 210L

Air Blank:
07/01/21 10:07 .000
Calibration Check:
22 07/01/21 10:07 .097

Subject Name

Test #2

Subject I.D.

N/A

Operator Name, I.D.

M. Nelson #200210

Location

Clinton P.D.

101 E. Ohio

Clinton, MO 64735

AS IV Serial no: 105448
Version no: 532B

TEST RECORD 01476

Temp Date Time 210L

Air Blank:
07/01/21 10:09 .000
Calibration Check:
23 07/01/21 10:09 .097

Subject Name

Test #3

Subject I.D.

N/A

Operator Name, I.D.

M. Nelson #200210

Location

Clinton P.D.

101 E. Ohio

Clinton, MO 64735

AS IV Serial no: 105448
Version no: 532B

TEST RECORD 01477

Temp Date Time 210L

VOID: RFI
12 07/01/21 10:11

Subject Name

RFI

Subject I.D.

N/A

Operator Name, I.D.

M. Nelson #200210

Location

Clinton P.D.

101 E. Ohio

Clinton, MO 64735

AS IV Serial no: 105448
Version no: 532B

TEST RECORD 01478

Temp Date Time 210L

Air Blank:
07/01/21 10:14 .000
Subject Test: RFI
23 07/01/21 10:14 .000

Subject Name

Self Test

Subject I.D.

N/A

Operator Name, I.D.

M. Nelson #200210

Location

Clinton P.D.

101 E. Ohio

Clinton, MO 64735



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 20190 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on April 8, 2020, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1206% (w/vol) ethyl alcohol. The expiration date for this lot number is April 6, 2022 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN02271802 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

MICHAEL S NELSON

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/17/2020

NUMBER 200210

EXPIRES 7/17/2022

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator NELSON, MICHAEL
Permit No 200210
Date Issued 7/17/2020 **Date Expires** 7/17/2022