

AS IV Serial no: 187460
Version no: 532B

TEST RECORD 88861

Temp Date Time 216L

Air Blank:
04/16/71 28:37 .000
Calibration Check:
04/16/71 28:37 .181

Subject Name
Test 1
Subject I.D.
M. Blake Montgomery
200254

Operator Name, I.D.
MPD
Location
461 W Arrow

.101

AS IV Serial no: 187460
Version no: 532B

TEST RECORD 88862

Temp Date Time 216L

Air Blank:
04/16/71 28:38 .000
Calibration Check:
04/16/71 28:38 .184

Subject Name
Test 2
Subject I.D.

Operator Name, I.D.
M Blake Montgomery
200254

Location
MPD
461 W Arrow
.104

AS IV Serial no: 187460
Version no: 532B

TEST RECORD 88863

Temp Date Time 216L

Air Blank:
04/16/71 28:48 .000
Calibration Check:
04/16/71 28:48 .182

Subject Name
Test 3
Subject I.D.

Operator Name, I.D.
M. Blake Montgomery
200254

Location
MPD
461 W. Arrow
.102

AS IV Serial no: 187460
Version no: 532B

TEST RECORD 88864

Temp Date Time 216L

VOID: RFI
12 04/16/71 28:46

Subject Name
Test RFI
Subject I.D.
M. Blake Montgomery
Operator Name, I.D.
200254

Location
MPD
461 W Arrow
RFI



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **20420** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **September 23, 2020**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1208%** (w/vol) ethyl alcohol. The expiration date for this lot number is **September 22, 2022** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03052002 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

**PERMIT
 TYPE II**

M. BLAKE MONTGOMERY

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 9/24/2020

NUMBER 200254

EXPIRES 9/24/2022

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator MONTGOMERY, M.
Permit No 200254
Date Issued 9/24/2020 **Date Expires** 9/24/2022