



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 107993	PRINTER SN 099.3586.836	DATE OF INSPECTION 10/09/2021
LOCATION OF INSTRUMENT (STREET AND CITY) 9623 SAINT CHARLES ROCK ROAD, BRECKENRIDGE HILLS POLICE DEPARTMENT		TIME OF INSPECTION 10:15 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER GUTH LABORATORIES, INC. LOT # 21080 EXP. DATE 03/08/2023
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN SD2309 SIMULATOR EXP DATE 03/04/2022

- CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .096	TEST 2 ← .096	TEST 3 ← .095
---------------	---------------	---------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE 	PRINT NAME RANDY VONCLOEDT
TYPE II PERMIT NUMBER/EXPIRATION DATE 280205 05/11/2022	TELEPHONE NUMBER (314) 426-1214

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AS IV Serial no: 187993
Version no: 8328

TEST RECORD 08090

Temp Date Time 2102

Air Blank:
18/09/21 22:17 .000
Calibration Check:
21 18/09/21 22:17 .016

Subject Name

Maint.

Subject I.D.

N/A

Operator Name, I.D.

Vn C 207

Location

9623 M0180

St. Louis MO 63114

Vn 207

AS IV Serial no: 187993
Version no: 8328

TEST RECORD 08091

Temp Date Time 2102

Air Blank:
18/09/21 22:18 .000
Calibration Check:
21 18/09/21 22:18 .016

Subject Name

Maint.

Subject I.D.

N/A

Operator Name, I.D.

Vn C 207

Location

9623 M0180

St. Louis MO 63114

Vn 207

AS IV Serial no: 187993
Version no: 8328

TEST RECORD 08092

Temp Date Time 2102

Air Blank:
18/09/21 22:20 .000
Calibration Check:
22 18/09/21 22:19 .016

Subject Name

Maint.

Subject I.D.

N/A

Operator Name, I.D.

Vn C 207

Location

9623 M0180

St. Louis MO 63114

Vn 207

AS IV Serial no: 187993
Version no: 8328

TEST RECORD 08093

Temp Date Time 2102

VOID: 12/18/21 22:21

Subject Name

Maint.

Subject I.D.

Operator Name, I.D.

Vn C 207

Location

9623 M0180

St. Louis MO 63114

Vn 207



GUTH LABORATORIES, INC.

500 NORTH 67th STREET • HARRISBURG, PA 17111-4811 • TELEPHONE: 717-564-9470


CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 21080 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on March 10, 2021, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1214% (w/vol) ethyl alcohol. The expiration date for this lot number is March 8, 2023 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- 2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.


Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:
Testing was conducted using Cerilliant Reference Standard lot number FN03032002 whose values are traceable to NIST.
All balances are calibrated annually by an outside agency using NIST traceable weights.
Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

RANDY J VON CLOEDT

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/11/2020
NUMBER 200171
EXPIRES 5/11/2022

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580,0771 (6-10)

LAB-4 (PG-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcohol content in breath from of expired air in Missouri.

Operator VON CLOEDT, RANDY
Permit No 200171
Date Issued 5/11/2020 Date Expires 5/11/2022