

AS IV Serial no: 107996
Version no: 532B

TEST RECORD 01277

| Temp | Date | Time | a/ 210L |
|------|------|------|------------|
|------|------|------|------------|

Air Blank:
03/04/21 21:55 .000
Calibration Check:
22 03/04/21 21:55 .100

Subject Name

TEST 1

Subject I.D.

Operator Name, I.D.

E. Stock 200077

Location

GLADSTONE PD

AS IV Serial no: 107996
Version no: 532B

TEST RECORD 01278

| Temp | Date | Time | a/ 210L |
|------|------|------|------------|
|------|------|------|------------|

Air Blank:
03/04/21 21:56 .000
Subject Test: Man
23 03/04/21 21:56 .100

Subject Name

TEST 2

Subject I.D.

Operator Name, I.D.

E. Stock 200077

Location

GLADSTONE PD

AS IV Serial no: 107996
Version no: 532B

TEST RECORD 01279

| Temp | Date | Time | a/ 210L |
|------|------|------|------------|
|------|------|------|------------|

Air Blank:
03/04/21 21:59 .000
Calibration Check:
23 03/04/21 21:59 .100

Subject Name

TEST 3

Subject I.D.

Operator Name, I.D.

E. Stock 200077

Location

GLADSTONE PD

AS IV Serial no: 107996
Version no: 532B

TEST RECORD 01280

| Temp | Date | Time | a/ 210L |
|------|------|------|------------|
|------|------|------|------------|

VOID: RFI
12 03/04/21 22:01

Subject Name

RFI TEST

Subject I.D.

Operator Name, I.D.

E. Stock 200077

Location

GLADSTONE PD

AS IV Serial no: 107996
Version no: 532B

TEST RECORD 01281

| Temp | Date | Time | a/ 210L |
|------|------|------|------------|
|------|------|------|------------|

Air Blank:
03/04/21 22:02 .000
Subject Test: Auto
24 03/04/21 22:02 .000

Subject Name

Blank TEST

Subject I.D.

Operator Name, I.D.

E. Stock 200077

Location

GLADSTONE PD



Airgas USA LLC (LAB)
3500 Bernard Street
St. Louis, Mo. 63103
Ph: (314) 533-3100
Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Exclusive Supplier
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 26-Mar-2019

Lot # AG908401 **Model** 108cadd

| <u>Exp. Date</u> | <u>Cyl. Type</u> | <u>Component</u> | <u>Certified Concentration</u> |
|------------------|------------------|---------------------|--------------------------------------|
| 25-Mar-2021 | 108 | Ethanol Nitrogen | 0.100 ± 2% BrAC (272 ppm) Balance |

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

| <u>RGM Serial No.</u> | <u>Concentration</u> | <u>RGM Serial No.</u> | <u>Concentration</u> |
|-----------------------|----------------------|-----------------------|----------------------|
| EB0010581 | 392.1 ppm | EB0010603 | 393.0 ppm |
| EB0010570 | 259.8 ppm | EB0010559 | 258.2 ppm |
| EB0010285 | 208.0 ppm | EB0010595 | 208.3 ppm |
| EB0010561 | 103.6 ppm | EB0010562 | 104.2 ppm |
| EB0010681 | 52.12 ppm | EB0010579 | 52.81 ppm |

| <u>CRM Serial No.</u> | <u>Concentration</u> | <u>CRM Serial No.</u> | <u>Concentration</u> |
|-----------------------|----------------------|-----------------------|----------------------|
| CC434668 | 800.0 ppm | 0056649 | 390.1 ppm |
| CC234503 | 253.0 ppm | 0056662 | 150.2 ppm |

Analytical Method: NDIR

Digitally signed by Quality Control
Date: 2019.03.26 12:19:36 -05:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Approved for Release: _____


Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07