



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 108263	NAME OF AGENCY WOODSON TERRACE	DATE OF INSPECTION 12/8/2021
LOCATION OF INSTRUMENT (STREET AND CITY) 4305 WOODSON RD, WOODSON TERRACE, MO		TIME OF INSPECTION 1330

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER **Guth** LOT # **21190** EXP. DATE **6/8/23**
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) **34.0** SIM. SN **DR3942** SIM. NIST EXP DATE **8/23/22**

- CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 • .102	TEST 2 • .102	TEST 3 • .102
----------------------	----------------------	----------------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 8	(0-.04) 8	(.05-.09) 8	(.10-.14) 8	(.15-.19) 1	(OVER .19) 8
-------------------	------------------	--------------------	--------------------	--------------------	---------------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

DEPARTMENT - WIDE TRAINING NOV. 30th AND DEC 7th. 120 TOTAL TESTS.

INSPECTING OFFICER

SIGNATURE 	PRINT NAME Anthony Truitt
TYPE II PERMIT NUMBER/EXPIRATION DATE 210169 8/4/23	TELEPHONE NUMBER 3149145978

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 108263
Version no: 532B

TEST RECORD 00601

Temp Date Time ^{s/} 210L

Air Blank:
12/08/21 13:46 .000
Calibration Check:
18 12/08/21 13:46 .102

Subject Name

THWEATT

Subject I.D.

SELF 210169

Operator Name, I.D.

SAME

Location

WTPD

AS IV Serial no: 108263
Version no: 532B

TEST RECORD 00602

Temp Date Time ^{s/} 210L

Air Blank:
12/08/21 13:48 .000
Calibration Check:
19 12/08/21 13:48 .102

Subject Name

THWEATT

Subject I.D.

SELF 210169

Operator Name, I.D.

SAME

Location

WTPD

AS IV Serial no: 108263
Version no: 532B

TEST RECORD 00603

Temp Date Time ^{s/} 210L

Air Blank:
12/08/21 13:50 .000
Calibration Check:
20 12/08/21 13:50 .102

Subject Name

THWEATT

Subject I.D.

SELF 210169

Operator Name, I.D.

SAME

Location

WTPD

AS IV Serial no: 108263
Version no: 532B

TEST RECORD 00604

Temp Date Time ^{s/} 210L

VOID: RFI
12 12/08/21 13:51

Subject Name

THWEATT

Subject I.D.

SELF 210169

Operator Name, I.D.

SAME

Location

WTPD



GUTH LABORATORIES, INC.

690 NORTH 67th STREET • HARRISBURG, PA 17111-4811 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **21190** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **June 9, 2021**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1206%** (w/vol) ethyl alcohol. The expiration date for this lot number is **June 8, 2023** at 11:59 PM.

When used in a calibrated Simulator, operating at **34°C +/- .2°C**, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

**Ted L. Pauley, President
GUTH LABORATORIES, INC.**

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03052002 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



Missouri Department of Health and Senior Services

P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010
RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466

Donald G. Kauerauf
Director



Michael L. Parson
Governor

SIMULATOR CERTIFICATION REPORT

SIMULATOR INFORMATION

Simulator Serial Number: DR3942 **Manufacturer:** Guth
Model Number: 2100
Agency: WOODSON TERRACE PD
Agency Address: 4305 WOODSON RD, WOODSON TERRACE, MO 63134

NIST THERMOMETER INFORMATION

Serial Number: 17KMM00690 **Bias:** 0.00
Uncertainty: 0.02
Date of Certification: 11/6/2020 **Date of Expiration:** 11/6/2021

ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

VERIFICATION RESULTS

<u>Simulator Average</u>	<u>NIST Average</u>	<u>Combined Uncertainty</u>
34.00	34.02	.04

The combined uncertainty is calculated with a k=2 value.

ADJUSTMENT RESULTS

No adjustment was needed.

Date of testing: 8/23/2021
Certification Expiration: 8/23/2022
Simulator testing technician: M. BOND

Notes on Condition: none

Deviation(s) from method: none

DHSS BAP Scientist Approving: B. LUTMER
Certification No: DR3942_8232021



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



**PERMIT
TYPE II**

ANTHONY THWEATT

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/4/2021

NUMBER 210169

EXPIRES 8/4/2023

Laura Q. Way

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MD 580-0771 (5-10)

LAB-4 (R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator THWEATT, ANTHONY
Permit No 210169
Date Issued 8/4/2021 **Date Expires** 8/4/2023

