



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 108264	NAME OF AGENCY Buckner PD	DATE OF INSPECTION 06/22/2021
LOCATION OF INSTRUMENT (STREET AND CITY) 315 S. Hudson St. Buckner, Mo 64016		TIME OF INSPECTION 9:59 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION

COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Intoxometers LOT # AG031402 EXP. DATE 07/09/2022

SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIM. SN SIM. NIST EXP DATE

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .098	TEST 2 .099	TEST 3 .098
--------------	--------------	--------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE 	PRINT NAME R. Kendall #316
TYPE II PERMIT NUMBER/EXPIRATION DATE 200199	TELEPHONE NUMBER (816) 650-3939

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



**PERMIT
 TYPE II**

RICHARD L. KENDALL

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/8/2020

NUMBER 200199

EXPIRES 7/8/2022

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 680-0771 (8-10)

LAB-4 (R6-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator KENDALL, RICHARD
Permit No 200199
Date Issued 7/8/2020 **Date Expires** 7/8/2022





Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63168
 Ph: (314) 533-3100
 Fax: (314) 533-7326

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 10-Nov-2020

Lot # AG031402 Model 34caod

Exp. Date 9-Jul-2022	Cyl. Type 34	Component Ethanol Nitrogen	Certified Concentration 0.100 ± 2% EtAC (272 ppm) Balance
--------------------------------	------------------------	---	--

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	292.1 ppm	EB0010683	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	268.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010561	52.12 ppm	EB0010579	52.81 ppm

<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
 Date: 2020.11.10 18:43:41 -0600
 Reason: Ety gas standard certification of analysis
 Location: Airgas USA LLC (Lab)

Approved for Release: _____

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07

AS IV Serial no: 108264
Version no: 532B

TEST RECORD 00555
Temp Date Time 210L s/
Air Blank: 06/22/21 21:59 .000
Calibration Check: 22 06/22/21 21:59 .098

Subject Name

MAINTENANCE

Subject I.D.

TEST 1

Operator Name, I.D.

KENDALL 200199

Location

AS IV Serial no: 108264
Version no: 532B

TEST RECORD 00557
Temp Date Time 210L s/
Air Blank: 06/22/21 22:04 .000
Calibration Check: 24 06/22/21 22:04 .098

Subject Name

MAINTENANCE

Subject I.D.

TEST 4

Operator Name, I.D.

KENDALL 200199

Location

AS IV Serial no: 108264
Version no: 532B

TEST RECORD 00556
Temp Date Time 210L s/
Air Blank: 06/22/21 22:01 .000
Calibration Check: 22 06/22/21 22:01 .099

Subject Name

MAINTENANCE

Subject I.D.

TEST 2

Operator Name, I.D.

KENDALL 200199

Location

AS IV Serial no: 108264
Version no: 532B

TEST RECORD 00558
Temp Date Time 210L s/
VOID: RFI
12 06/22/21 22:10

Subject Name

RFI

Subject I.D.

RFI

Operator Name, I.D.

KENDALL 200199

Location

AS IV Serial no: 108264
Version no: 532B

TEST RECORD 00556
Temp Date Time 210L s/
Air Blank: 06/22/21 22:02 .000
Calibration Check: 23 06/22/21 22:02 .098

Subject Name

MAINTENANCE

Subject I.D.

TEST 3

Operator Name, I.D.

KENDALL 200199

Location

AS IV Serial no: 108264
Version no: 532B

TEST RECORD 00559
Temp Date Time 210L s/
Air Blank: 06/22/21 22:12 .000
Subject Test: Auto
24 06/22/21 22:12 .000

Subject Name

SUBJECT

Subject I.D.

TEST 1

Operator Name, I.D.

KENDALL 200199

Location