



ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 108269	PRINTER SN 099.3586.809	DATE OF INSPECTION 11/03/2021
-----------------------------	----------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) Zone D-21 / Kimberling City, MO	TIME OF INSPECTION 10:20 am
---	--------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

<input checked="" type="checkbox"/> SIMULATOR SOLUTION	<input type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>RepCo Marketing, Inc.</u> LOT # <u>20001</u> EXP. DATE <u>10/07/2022</u>	
<input checked="" type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C) <u>34.00</u> SIMULATOR SN <u>MP2307</u> SIMULATOR EXP DATE <u>01/10/2020</u>	

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← 0.100	TEST 2 ← 0.100	TEST 3 ← 0.101
----------------	----------------	----------------

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Calibration performed.

INSPECTING OFFICER	
SIGNATURE <i>DR Hukill #383</i>	PRINT NAME Tpr. D.R. Hukill / #383
TYPE II PERMIT NUMBER/EXPIRATION DATE 210065 / 04/06/2023	TELEPHONE NUMBER (417) 895-6868

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AS IV Serial no: 108269
Version no: 532B

TEST RECORD 00582
s/
Temp Date Time 210L

VOID: RFI
12 11/03/21 11:11

Subject Name
Accuracy Check
Subject I.D.
RFI Test
Operator Name, I.D.
Tpr. D.R. Hukill #383
Location
Zone D-21
Kimberling City

AS IV Serial no: 108269
Version no: 532B

TEST RECORD 00581
s/
Temp Date Time 210L

Air Blank:
11/03/21 11:08 .000
Calibration Check:
26 11/03/21 11:08 .101

Subject Name
Accuracy Check
Subject I.D.
Test #3
Operator Name, I.D.
Tpr. D.R. Hukill #383
Location
Zone D-21
Kimberling City

AS IV Serial no: 108269
Version no: 532B

TEST RECORD 00580
s/
Temp Date Time 210L

Air Blank:
11/03/21 11:06 .000
Calibration Check:
26 11/03/21 11:06 .100

Subject Name
Accuracy Check
Subject I.D.
Test #2
Operator Name, I.D.
Tpr. D.R. Hukill #383
Location
Zone D-21
Kimberling City

AS IV Serial no: 108269
Version no: 532B

TEST RECORD 00579
s/
Temp Date Time 210L

Air Blank:
11/03/21 11:03 .000
Calibration Check:
25 11/03/21 11:03 .100

Subject Name
Accuracy Check
Subject I.D.
Test #1
Operator Name, I.D.
Tpr. D.R. Hukill #383
Location
Zone D-21
Kimberling City

AS IV Serial no: 108269
Version no: 532B

TEST RECORD 00578
s/
Temp Date Time 210L

Air Blank:
11/03/21 11:01 .000
Calibration:
24 11/03/21 11:01 .100

Subject Name
Calibration
Subject I.D.
Test
Operator Name, I.D.
Tpr. D.R. Hukill #383
Location
Zone D-21
Kimberling City



RepCo Marketing Co
3101-188 Stony Brook Dri
Raleigh, NC 27604
919-876-5480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing Co.
LOT NUMBER: 20001
EXPIRATION DATE: October 7, 2022 at 11:59 p.m.

RepCo Marketing Co. certifies the following:


RepCo Marketing Co. prepared, tested and supplied Lot Number 20001 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by ISO 17025 and ISO 17034 accredited institution Alcohol Countermeasure Systems, using NIST standards. Random samples were analyzed by Alcohol Countermeasure Systems utilizing a gas chromatograph and found to contain .1227 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is October 8, 2020 The expiration date for this lot number is October 7, 2022 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.


Alma Palmer, Operations Manager
RepCo Marketing Co.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
DARWIN R. HUKILL

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/6/2021

NUMBER 210065

EXPIRES 4/6/2023

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator HUKILL, DARWIN
 Permit No 210065
 Date Issued 4/6/2021 Date Expires 4/6/2023

