



AS IV Serial no: 111320  
Version no: 532B

TEST RECORD 00427

Temp Date Time <sup>a/</sup> 210L

Air Blank:  
04/15/21 13:30 .000  
Calibration Check:  
23 04/15/21 13:30 .101

Subject Name

Test #1

Subject I.D.

N/A

Operator Name, I.D.

PO Newman #3072

Location

MODOT TMC

AS IV Serial no: 111320  
Version no: 532B

TEST RECORD 00428

Temp Date Time <sup>a/</sup> 210L

Air Blank:  
04/15/21 13:33 .000  
Calibration Check:  
24 04/15/21 13:33 .099

Subject Name

Test #2

Subject I.D.

N/A

Operator Name, I.D.

PO Newman #3072

Location

MODOT TMC

AS IV Serial no: 111320  
Version no: 532B

TEST RECORD 00429

Temp Date Time <sup>a/</sup> 210L

Air Blank:  
04/15/21 13:35 .000  
Calibration Check:  
25 04/15/21 13:35 .099

Subject Name

Test #3

Subject I.D.

N/A

Operator Name, I.D.

PO Newman 3072

Location

MODOT TMC

AS IV Serial no: 111320  
Version no: 532B

TEST RECORD 00430

Temp Date Time <sup>a/</sup> 210L

VOID: RFI  
12 04/15/21 13:37

Subject Name

RFF

Subject I.D.

N/A

Operator Name, I.D.

PO Newman 3072

Location

MODOT TMC



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**

*Exclusive Supplier*  
Intoximeters, Inc.  
2081 Craig Road  
St. Louis, Mo 63146

**Test Date:** 8-Jun-2020

**Lot # AG015503 Model 108cacc**

**Exp. Date**

3-Jun-2022

**Cyl. Type**

108

**Component**

Ethanol  
Nitrogen

**Certified Concentration**

0.100 ± 2% BrAC (272 ppm)  
Balance

**Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:**

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm
<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC434668	800.0 ppm	0056649	390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

**Analytical Method:** NDIR

Digitally signed by Quality Control  
Date: 2020.06.10 14:11:39 -05:00  
Reason: Dry gas standard certification of analysis  
Location: Airgas USA LLC (Lab)

Approved for Release:

Rod Marsala

**ISO 17025:2005 A2LA accredited. Certificate Number 3082.06**  
**ISO 17034:2016 A2LA accredited. Certificate Number 3082.07**



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT  
 TYPE II**

**DAWN M. NEUMAN**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX DMT**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/6/2021

NUMBER 210074

EXPIRES 4/6/2023

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES