



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
 By Tracy Crews at 10:25 am, Jun 08, 2021

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111326	PRINTER SN 099.3586.615	DATE OF INSPECTION 06/07/2021
LOCATION OF INSTRUMENT (STREET AND CITY) Richland PD 201 South Chestnut Street Richland Missouri 65556		TIME OF INSPECTION 3:24 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION
- COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER Intoximeters LOT # AG109702 EXP. DATE 04/07/2023

SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIMULATOR SN SIMULATOR EXP DATE

- CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← 0.102	TEST 2 ← 0.099	TEST 3 ← 0.099
----------------	----------------	----------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Instrument is operating within established limits.

INSPECTING OFFICER		PRINT NAME
SIGNATURE	<i>David L. Moser</i>	David L. Moser
TYPE II PERMIT NUMBER/EXPIRATION DATE		TELEPHONE NUMBER
200277 11/05/2022		(573) 480-4144

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901

JUN/07/2021/MON 04:55 PM Richland PD FAX No. 573 765 3093 P. 002

AS IV Serial no: 111326
Version no: 532B

TEST RECORD 00791

Temp Date Time 210L

Air Blank:
06/07/21 15:24 .000
Calibration Check:
22 06/07/21 15:24 .000

Subject Name

BLANK CHECK
Subject I.D.

Operator Name: I.D.

DMOSER 200277
Location
RICHMOND RD

AS IV Serial no: 111326
Version no: 532B

TEST RECORD 00793

Temp Date Time 210L

Air Blank:
06/07/21 15:28 .000
Subject Test: Man
23 06/07/21 15:28 .102

Subject Name

TEST 1
Subject I.D.

Operator Name: I.D.

DMOSER 200277
Location
RICHMOND RD

AS IV Serial no: 111326
Version no: 532B

TEST RECORD 00794

Temp Date Time 210L

Air Blank:
06/07/21 15:30 .000
Subject Test: Man
23 06/07/21 15:30 .099

Subject Name

TEST 2
Subject I.D.

Operator Name: I.D.

DMOSER 200277
Location
RICHMOND RD

AS IV Serial no: 111326
Version no: 532B

TEST RECORD 00795

Temp Date Time 210L

Air Blank:
06/07/21 15:32 .000
Subject Test: Man
23 06/07/21 15:32 .099

Subject Name

TEST 3
Subject I.D.

Operator Name: I.D.

DMOSER 200277
Location
RICHMOND RD

AS IV Serial no: 111326
Version no: 532B

TEST RECORD 00796

Temp Date Time 210L

VOID: RFI
12 06/07/21 15:33

Subject Name

RFI TEST
Subject I.D.

Operator Name: I.D.

DMOSER 200277
Location
RICHMOND RD



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 8-Apr-2021

Lot # AG109702 Model 108cadd

<u>Exp. Date</u> 7-Apr-2023	<u>Cyl. Type</u> 108	<u>Component</u> Ethanol Nitrogen	<u>Certified Concentration</u> 0.100 ± 2% BrAC (272 ppm) Balance
--------------------------------	-------------------------	---	--

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
 Date: 2021.04.08 18:24:29 -05:00
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)

Approved for Release:

 Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
DAVID L. MOSER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 11/5/2020

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 200277

EXPIRES 11/5/2022

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 587-0771 (6-10)

LAB-4 (R8-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator MOSER, DAVID
Permit No 200277
Date issued 11/5/2020 Date Expires 11/5/2022