

AS IV Serial no: 111334
Version no: 532B

TEST RECORD 00309

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
07/03/21 18:06 .000
Calibration Check:
26 07/03/21 18:06 .096

Subject Name

Subject I.D.

Operator Name, I.D.

B. DS 9004
Location
Joplin PD

AS IV Serial no: 111334
Version no: 532B

TEST RECORD 00310

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
07/03/21 18:07 .000
Calibration Check:
26 07/03/21 18:07 .097

Subject Name

Subject I.D.

Operator Name, I.D.

B. DS 9004
Location
Joplin PD

AS IV Serial no: 111334
Version no: 532B

TEST RECORD 00311

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
07/03/21 18:09 .000
Calibration Check:
27 07/03/21 18:09 .097

Subject Name

Subject I.D.

Operator Name, I.D.

B. DS 9004
Location
Joplin PD

AS IV Serial no: 111334
Version no: 532B

TEST RECORD 00312

Temp	Date	Time	s/ 210L
------	------	------	------------

VOID: RFI
12 07/03/21 18:11

Subject Name

Subject I.D.

Operator Name, I.D.

B. DS 9004
Location
Joplin PD



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

Certificate of Analysis

Customer Name

Exclusive Supplier

Intoximeters, Inc.

2081 Craig Road

St. Louis, Mo 63146

Test Date: 26-May-2020

Lot # AG014303 Model 55cacc

Exp. Date

22-May-2022

Cyl. Type

55

Component

Ethanol

Nitrogen

Certified Concentration

0.100 ± 2% BrAC (272 ppm)

Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.

EB0010581

EB0010570

EB0010285

EB0010561

EB0010681

Concentration

392.1 ppm

259.8 ppm

208.0 ppm

103.6 ppm

52.12 ppm

RGM Serial No.

EB0010603

EB0010559

EB0010595

EB0010562

EB0010579

Concentration

393.0 ppm

258.2 ppm

208.3 ppm

104.2 ppm

52.81 ppm

CRM Serial No.

CC434668

CC234503

Concentration

800.0 ppm

253.0 ppm

CRM Serial No.

0056649

0056662

Concentration

390.1 ppm

150.2 ppm

Analytical Method:

NDIR

Digitally signed by Quality Control
Date: 2020.05.28 14:58:35 -05:00
Reason: Dry gas standard certification of analysts
Location: Airgas USA LLC (Lab)

Approved for Release:



Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06

ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
BRETT DAVIS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 10/11/2019

NUMBER 290242

EXPIRES 10/11/2021

MO 580-0771 (5-10)


 DIRECTOR OF STATE PUBLIC HEALTH LABORATORY


 DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R5-10)

 **STATE OF MISSOURI**
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator **DAVIS, BRETT**
 Permit No **290242**
 Date Issued **10/11/2019** Date Expires **10/11/2021**

