



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
 By Tracy Crews at 11:58 am, Jul 02, 2021

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111629	NAME OF AGENCY University of Missouri Police Department	DATE OF INSPECTION 07/02/2021
LOCATION OF INSTRUMENT (STREET AND CITY) 901 Virginia Avenue, Columbia 65211		TIME OF INSPECTION 7:35 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER Intoximeters LOT # AG021102 EXP. DATE 07/29/2022
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIM. SN _____ SIM. NIST EXP DATE _____

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .078	TEST 2 .078	TEST 3 .078
--------------	--------------	--------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	1	(.05-.09)	4	(.10-.14)	1	(.15-.19)	2	(OVER .19)	1
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE 	PRINT NAME Eric Moss
TYPE II PERMIT NUMBER/EXPIRATION DATE 200215 07/23/2021	TELEPHONE NUMBER (573) 882-7201

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

NS IV Serial no. 111629
Version no. 532B

TEST RECORD 01829

Temp Date Time ^{o/} 219L

Air Blank:
07/02/21 07:35 .000
Calibration Check:
24 07/02/21 07:35 .077

Subject Name

Cal Check 1 of 3 OK
Subject I.D.

Eric Moss MUPD 37
Operator Name, I.D.

200215 7/23/22
Location

MUPD

NS IV Serial no. 111629
Version no. 532B

TEST RECORD 01830

Temp Date Time ^{o/} 219L

Air Blank:
07/02/21 07:36 .000
Calibration Check:
25 07/02/21 07:36 .078

Subject Name

Cal Check 2 of 3 OK
Subject I.D.

Eric Moss MUPD 37
Operator Name, I.D.

200215 7/23/22
Location

MUPD

NS IV Serial no. 111629
Version no. 532B

TEST RECORD 01831

Temp Date Time ^{o/} 219L

Air Blank:
07/02/21 07:38 .000
Calibration Check:
25 07/02/21 07:38 .078

Subject Name

Cal Check 2 of 3 OK
Subject I.D.

Eric Moss MUPD 37
Operator Name, I.D.

200215 7/23/22
Location

MUPD

NS IV Serial no. 111629
Version no. 532B

TEST RECORD 01832

Temp Date Time ^{o/} 219L

UNIT: RFI
12 07/02/21 07:39

Subject Name

RFI test OK
Subject I.D.

Eric Moss MUPD 37
Operator Name, I.D.

200215 7/23/22
Location

MUPD

NS IV Serial no. 111629
Version no. 532B

TEST RECORD 01833

Temp Date Time ^{o/} 219L

Air Blank:
07/02/21 07:40 .000
Subject Test: Auto
26 07/02/21 07:40 .000

Subject Name

Blank test OK
Subject I.D.

Eric Moss MUPD 37
Operator Name, I.D.

200215 7/23/22
Location

MUPD



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

Certificate of Analysis

Customer Name

Exclusive Supplier

Intoximeters, Inc.

2081 Craig Road

St. Louis, Mo 63146

Test Date: 4-Aug-2020

Lot # AG021102 **Model** 108cacc

Exp. Date

29-Jul-2022

Cyl. Type

108

Component

Ethanol

Nitrogen

Certified Concentration

0.080 ± 0.002 BrAC (218 ppm)

Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.

EB0010581

EB0010570

EB0010285

EB0010561

EB0010681

Concentration

392.1 ppm

259.8 ppm

208.0 ppm

103.6 ppm

52.12 ppm

RGM Serial No.

EB0010603

EB0010559

EB0010595

EB0010562

EB0010579

Concentration

393.0 ppm

258.2 ppm

208.3 ppm

104.2 ppm

52.81 ppm

CRM Serial No.

CC434668

CC234503

Concentration

800.0 ppm

253.0 ppm

CRM Serial No.

0056649

0056662

Concentration

390.1 ppm

150.2 ppm

Analytical Method:

NDIR

Digitally signed by Quality Control
Date: 2020.08.06 19:18:53 -05:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Approved for Release:

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06

ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
ERIC L MOSS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/23/2020

NUMBER 200215

EXPIRES 7/23/2022

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator MOSS, ERIC
 Permit No 200215
 Date Issued 7/23/2020 Date Expires 7/23/2022

