



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
by Tracy Green at 8:54 am, Aug 27, 2021

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

| | | |
|--|-------------------------------|----------------------------------|
| ALCO SENSOR IV SN 111647 | NAME OF AGENCY NORMANDY PD | DATE OF INSPECTION 08/26/2021 |
| LOCATION OF INSTRUMENT (STREET AND CITY) 7700 NATURAL BRIDGE, NORMANDY, MO, 63121 | | TIME OF INSPECTION 8:54 am |

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER GUTH LOT # 21080 EXP. DATE 03/08/2023

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIM. SN SD1615 SIM. NIST EXP DATE 07/28/2022

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .101

TEST 2 .101

TEST 3 .101

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

| | | | | | |
|----------|---------|-----------|-------------|-------------|------------|
| REFUSALS | (0-.04) | (.05-.09) | (.10-.14) 1 | (.15-.19) 2 | (OVER .19) |
|----------|---------|-----------|-------------|-------------|------------|

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE *Brandon Lange*

PRINT NAME
BRANDON LANGE

TYPE II PERMIT NUMBER/EXPIRATION DATE
210148 07/23/2023

TELEPHONE NUMBER
(314) 385-3300

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111647
Version no: 532B

TEST RECORD 00330

Temp Date Time ^{s/} 210L

Air Blank:
08/26/21 08:54 .000
Calibration Check:
19 08/26/21 08:54 .101

Subject Name

maint 1

Subject I.D.

N/A

Operator Name, I.D.

Danay 266

Location

Normandy PD

AS IV Serial no: 111647
Version no: 532B

TEST RECORD 00331

Temp Date Time ^{s/} 210L

Air Blank:
08/26/21 08:55 .000
Calibration Check:
19 08/26/21 08:55 .101

Subject Name

maint 2

Subject I.D.

N/A

Operator Name, I.D.

Danay 266

Location

Normandy PD

AS IV Serial no: 111647
Version no: 532B

TEST RECORD 00332

Temp Date Time ^{s/} 210L

Air Blank:
08/26/21 08:57 .000
Calibration Check:
20 08/26/21 08:57 .101

Subject Name

maint 3

Subject I.D.

N/A

Operator Name, I.D.

Danay 266

Location

Normandy PD

AS IV Serial no: 111647
Version no: 532B

TEST RECORD 00333

Temp Date Time ^{s/} 210L

VOID: RFI
12 08/26/21 08:58

Subject Name

R.F.I

Subject I.D.

N/A

Operator Name, I.D.

Danay 266

Location

Normandy PD

AS IV Serial no: 111647
Version no: 532B

TEST RECORD 00334

Temp Date Time ^{s/} 210L

Air Blank:
08/26/21 09:00 .000
Subject Test: Auto
21 08/26/21 09:00 .000

Subject Name

Danay

Subject I.D.

266

Operator Name, I.D.

Her S

Location

Normandy PD



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **21080** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **March 10, 2021**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1214%** (w/vol) ethyl alcohol. The expiration date for this lot number is **March 8, 2023** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03052002 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
BRANDON W. LANGE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/23/2021

NUMBER 210148

EXPIRES 7/23/2023

Laura E. Wag

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Rob Knud

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator LANGE, BRANDON
 Permit No 210148
 Date Issued 7/23/2021 Date Expires 7/23/2023

