



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
 By Tracy Crews at 11:44 am, Jul 27, 2021

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111660	NAME OF AGENCY Fredericktown PD	DATE OF INSPECTION 07/26/2021
-----------------------------	------------------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 120 W. Main St. - Fredericktown	TIME OF INSPECTION 9:43 am
---	-------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER GUTH Laboratories LOT # 21080 EXP. DATE 02/02/2022

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIM. SN MP2936 SIM. NIST EXP DATE 02/22/2022

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .100

TEST 2 ← .099

TEST 3 ← .099

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Instrument is operating properly

INSPECTING OFFICER

SIGNATURE

PRINT NAME
Sgt. Michael Sletten

TYPE II PERMIT NUMBER/EXPIRATION DATE
200271 - 10/19/2022

TELEPHONE NUMBER
(573) 783-3660

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

**PERMIT
 TYPE II**

MICHAEL D SLETTEN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 10/19/2020

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 200271

EXPIRES 10/19/2022

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The permit cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator SLETTEN, MICHAEL
 Permit No 200271
 Date Issued 10/19/2020 Date Expires 10/19/2022



Missouri Department of Health and Senior Services
 P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010
 RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466
 Randall W. Williams, MD, FACOG
 Director



Michael L. Parson
 Governor

SIMULATOR CERTIFICATION REPORT

SIMULATOR INFORMATION

Simulator Serial Number: MP2936 Manufacturer: Guth
 Model Number: 12V500
 Agency: FREDERICKTOWN PD
 Agency Address: 120 W MAIN ST, FREDERICKTOWN, MO 63645

NIST THERMOMETER INFORMATION

Serial Number: 17KMM00690 Bias: 0.00
 Uncertainty: 0.02
 Date of Certification: 11/6/2020 Date of Expiration: 11/6/2021

ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

VERIFICATION RESULTS

<u>Simulator Average</u>	<u>NIST Average</u>	<u>Combined Uncertainty</u>
34.00	34.02	.04

The combined uncertainty is calculated with a k=2 value.

ADJUSTMENT RESULTS

No adjustment was needed.

Date of testing: 2/2/2021
 Certification Expiration: 2/2/2022
 Simulator testing technician: D. DEBOARD

Notes on Condition: none

Deviation(s) from method: none

DHSS BAP Scientist Approving: B. LUTMER
 Certification No: MP2936_222021

X

DHSS BAP Scientist Approving



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **21080** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **March 10, 2021**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1214%** (w/vol) ethyl alcohol. The expiration date for this lot number is **March 8, 2023** at 11:59 PM.

When used in a calibrated Simulator, operating at **34°C +/- .2°C**, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03052002 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

AS IV Serial no: 111660
Version no: 532B

TEST RECORD 00610

Temp Date Time ^{s/} 210L

Air Blank:
07/26/21 09:43 .000
Subject Test: Auto
20 07/26/21 09:43 .000

Subject Name

BLANK TEST

Subject I.D.

MONTHLY MAINT.

Operator Name, I.D.

M. SLETTEN - 200271

Location

FREDERICKTOWN PD

AS IV Serial no: 111660
Version no: 532B

TEST RECORD 00611

Temp Date Time ^{s/} 210L

Air Blank:
07/26/21 09:45 .000
Calibration Check:
21 07/26/21 09:45 .100

Subject Name

SAMPLE TEST #1

Subject I.D.

MONTHLY MAINT.

Operator Name, I.D.

M. SLETTEN - 200271

Location

FREDERICKTOWN PD

AS IV Serial no: 111660
Version no: 532B

TEST RECORD 00612

Temp Date Time ^{s/} 210L

Air Blank:
07/26/21 09:46 .000
Calibration Check:
22 07/26/21 09:46 .099

Subject Name

SAMPLE TEST #2

Subject I.D.

MONTHLY MAINT.

Operator Name, I.D.

M. SLETTEN - 200271

Location

FREDERICKTOWN PD

AS IV Serial no: 111660
Version no: 532B

TEST RECORD 00613

Temp Date Time ^{s/} 210L

Air Blank:
07/26/21 09:48 .000
Calibration Check:
22 07/26/21 09:48 .099

Subject Name

SAMPLE TEST #3

Subject I.D.

MONTHLY MAINT.

Operator Name, I.D.

M. SLETTEN - 200271

Location

FREDERICKTOWN PD

AS IV Serial no: 111660
Version no: 532B

TEST RECORD 00614

Temp Date Time ^{s/} 210L

VOID: RFI
12 07/26/21 09:49

Subject Name

RFI CHECK

Subject I.D.

MONTHLY MAINT.

Operator Name, I.D.

M. SLETTEN - 200271

Location

FREDERICKTOWN PD

AS IV Serial no: 111668
Version no: 532B

Last Calibration:
11/25/19 15:36 .100

Test Results:
00606 Subject Test: Man
-20 07/02/21 14:47 .142
01.2 Lit. 04.8 Sec.
00607 VOID: RFI
12 07/02/21 14:48
00.0 Lit. 00.0 Sec.
00608 Subject Test: Man
-21 07/02/21 14:49 .140
00.7 Lit. 03.3 Sec.
00609 Subject Test: Man
-21 07/02/21 14:51 .140
01.0 Lit. 04.0 Sec.
00610 Subject Test: Auto
-20 07/26/21 09:43 .000
01.0 Lit. 04.3 Sec.
00611 Calibration Check:
-21 07/26/21 09:45 .100
00.0 Lit. 00.0 Sec.
00612 Calibration Check:
-22 07/26/21 09:46 .099
00.0 Lit. 00.0 Sec.
00613 Calibration Check:
-22 07/26/21 09:48 .099
00.0 Lit. 00.0 Sec.
00614 VOID: RFI
12 07/26/21 09:49
00.0 Lit. 00.0 Sec.
