



Serial no: 111664  
Version no: 532B

TEST RECORD 00188

Date Time 210L

Blank:  
08/14/21 14:54 .000  
Position Check:  
08/14/21 14:54 .102

Name  
ANDREW LIARD

I.D.  
806

Operator Name, I.D.  
And Liard 806

Location  
ZONE 21 OFFICE

Activity  
MAINT CHECK

Serial no: 111664  
Version no: 532B

TEST RECORD 00187

Date Time 210L

Blank:  
08/14/21 14:51 .000  
Position Check:  
08/14/21 14:51 .103

Name  
ANDREW LIARD

I.D.  
806

Operator Name, I.D.  
And Liard 806

Location  
ZONE 21 OFFICE

Activity  
MAINT CHECK

AS IV Serial no: 111664  
Version no: 532B

TEST RECORD 00191

Date Time 210L

VOID: RFI  
12 08/14/21 15:09

Subject Name  
ANDREW LIARD

Subject I.D.  
806

Operator Name, I.D.  
And Liard 806

Location  
ZONE 21 OFFICE

Activity  
MAINT CHECK

Serial no: 111664  
Version no: 532B

TEST RECORD 00189

Date Time 210L

Blank:  
08/14/21 15:00 .000  
Position Check:  
08/14/21 15:00 .101

Name  
ANDREW LIARD

I.D.  
806

Operator Name, I.D.  
And Liard 806

Location  
ZONE 21 OFFICE

Activity  
MAINT CHECK

## CERTIFICATE OF ANALYSIS

**MANUFACTURER AND SUPPLIER: RepCo Marketing Co.**  
**LOT NUMBER: 19002**  
**EXPIRATION DATE: October 16, 2021 at 11:59 p.m.**

RepCo Marketing Co. certifies the following:


RepCo Marketing Co. prepared, tested and supplied Lot Number 19002 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by ISO 17025 and ISO 17034 accredited institution Alcohol Countermeasure Systems, using NIST standards. Random samples were analyzed by Alcohol Countermeasure Systems utilizing a gas chromatograph and found to contain .1231 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is October 17, 2019. The expiration date for this lot number is October 16, 2021 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.

  
Alma Palmer, Operations Manager  
RepCo Marketing Co.



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**ANDREW B. WARD**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**INTOX DMT**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE **12/30/2020**

NUMBER **200313**

EXPIRES **12/30/2022**

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

 **STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

**Operator** WARD, ANDREW  
**Permit No** 200313  
**Date Issued** 12/30/2020 **Date Expires** 12/30/2022

