



ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED

By Tracy Crews at 7:49 am, Nov 30, 2021

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111664	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 11/29/2021
-----------------------------	---	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) Kimberling City Zone 21 Office	TIME OF INSPECTION 9:16 am
--	-------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER RepCo LOT # 20001 EXP. DATE 10/07/2022

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34 SIM. SN MP2307 SIM. NIST EXP DATE 01/07/2022

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .099

TEST 2 .099

TEST 3 .099

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

This instrument is operating within Department of Health Standards.

Adjusted time for Daylight savings.

.10% solution used

INSPECTING OFFICER

SIGNATURE *Andrew B. Ward*

PRINT NAME
Andrew B. Ward

TYPE II PERMIT NUMBER/EXPIRATION DATE
200313 12/30/2022

TELEPHONE NUMBER
(417) 895-6868

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111664
Version no: 532B

TEST RECORD 00211

Temp	Date	Time	s/210L
Air Blank:			
	11/29/21	09:20	.000
Calibration Check:			
25	11/29/21	09:20	.099

Subject Name
ANDREW WARD
Subject I.D.
Andrew Ward
Operator Name, I.D.
ZONE 21 OFFICE
Location
MONTHLY MAINT
CHECK

AS IV Serial no: 111664
Version no: 532B

TEST RECORD 00212

Temp	Date	Time	s/210L
Air Blank:			
	11/29/21	09:26	.000
Calibration Check:			
24	11/29/21	09:26	.099

Subject Name
ANDREW WARD
Subject I.D.
Andrew Ward
Operator Name, I.D.
ZONE 21 OFFICE
Location
MONTHLY MAINT
CHECK

AS IV Serial no: 111664
Version no: 532B

TEST RECORD 00210

Temp	Date	Time	s/210L
Air Blank:			
	11/29/21	09:17	.000
Calibration Check:			
24	11/29/21	09:17	.099

Subject Name
ANDREW WARD
Subject I.D.
Andrew Ward
Operator Name, I.D.
ZONE 21 OFFICE
Location
MONTHLY MAINT
CHECK

AS IV Serial no: 111664
Version no: 532B

TEST RECORD 00213

Temp	Date	Time	s/210L	
VOID: RPI				
	12	11/29/21	09:28	

Subject Name
ANDREW WARD
Subject I.D.
Andrew Ward
Operator Name, I.D.
ZONE 21 OFFICE
Location
MONTHLY MAINT
CHECK



RepCo Marketing Co
3101-188 Stony Brook Drive
Raleigh, NC 27604
919-876-5480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing Co.
LOT NUMBER: 20001
EXPIRATION DATE: October 7, 2022 at 11:59 p.m.

RepCo Marketing Co. certifies the following:

RepCo Marketing Co. prepared, tested and supplied Lot Number 20001 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by ISO 17025 and ISO 17034 accredited institution Alcohol Countermeasure Systems, using NIST standards. Random samples were analyzed by Alcohol Countermeasure Systems utilizing a gas chromatograph and found to contain .1227 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is October 8, 2020 The expiration date for this lot number is October 7, 2022 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.

Alma Palmer, Operations Manager
RepCo Marketing Co.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
ANDREW B. WARD

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/30/2020

NUMBER 200313

EXPIRES 12/30/2022

Laura Q. Day

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Rob Knud

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

 **STATE OF MISSOURI**
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator WARD, ANDREW
 Permit No 200313
 Date Issued 12/30/2020 Date Expires 12/30/2022

