

AS IV Serial no: 111668
Version no: 532B

TEST RECORD 00018

Temp Date Time ^{9/} 210L

Air Blank:
05/11/21 13:54 .000
Calibration Check:
19 05/11/21 13:54 .104

Subject Name

TEST

Subject I.D.

#1

Operator Name, I.D.

MATT Dond 290214

Location 9-18-2021

MSC

AS IV Serial no: 111668
Version no: 532B

TEST RECORD 00019

Temp Date Time ^{9/} 210L

Air Blank:
05/11/21 13:56 .000
Calibration Check:
20 05/11/21 13:56 .102

Subject Name

TEST

Subject I.D.

#2

Operator Name, I.D.

MATT Dond 290214

Location 9-18-2021

MSC

AS IV Serial no: 111668
Version no: 532B

TEST RECORD 00020

Temp Date Time ^{9/} 210L

Air Blank:
05/11/21 13:57 .000
Calibration Check:
21 05/11/21 13:57 .102

Subject Name

TEST

Subject I.D.

#3

Operator Name, I.D.

MATT Dond 290214

Location 9-18-2021

MSC

AS IV Serial no: 111668
Version no: 532B

TEST RECORD 00021

Temp Date Time ^{9/} 210L

VOID: RFI
12 05/11/21 13:59

Subject Name

TEST

Subject I.D.

RFI

Operator Name, I.D.

MATT Dond 290214

Location 9-18-2021

MSC

AS IV Serial no: 111668
Version no: 532B

TEST RECORD 00022

Temp Date Time ^{9/} 210L

Air Blank:
05/11/21 14:00 .000
Subject Test: Auto
22 05/11/21 14:00 .000

Subject Name

Blank

Subject I.D.

TEST

Operator Name, I.D.

MATT Dond 290214

Location 9-18-2021

MSC



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **20190** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **April 8, 2020**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1206%** (w/vol) ethyl alcohol. The expiration date for this lot number is **April 6, 2022** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN02271802 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

**PERMIT
TYPE II**

MATT B BOND

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

INTOX DMT, INTOXILYZER 8000, INTOX EC/IR II, ASIIV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 9/19/2019

NUMBER 290214

EXPIRES 9/19/2021

[Signature]

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

[Signature]

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MC 580-0771 (6-10)

LAB-4, (RS-10)