



AS IV Serial no: 111670  
Version no: 532B

TEST RECORD 00247  
Temp Date Time 210L  
a/

Air Blank: 06/16/21 02:40 .000  
Calibration Check: 22 06/16/21 02:40 .081

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 111670  
Version no: 532B

TEST RECORD 00248  
Temp Date Time 210L  
a/

Air Blank: 06/16/21 02:41 .000  
Calibration Check: 23 06/16/21 02:41 .080

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 111670  
Version no: 532B

TEST RECORD 00249  
Temp Date Time 210L  
a/

Air Blank: 06/16/21 02:43 .000  
Calibration Check: 23 06/16/21 02:43 .081

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 111670  
Version no: 532B

TEST RECORD 00251

Temp Date Time 210L  
a/

VOID: RFI  
12 06/16/21 02:46

Subject Name

Subject I.D.

Operator Name, I.D.

Location



Airgas USA LLC (LAB)  
 3500 Bernard Street  
 St. Louis, Mo. 63103  
 Ph: (314) 533-3100  
 Fax: (314) 533-7328

## Certificate of Analysis

Customer Name  
 Exclusive Supplier  
 Intoximeters, Inc.  
 2081 Craig Road  
 St. Louis, Mo 63146

Test Date: 2-Oct-2018

**Lot # AG827002 Model 108caccd**

<u>Exp. Date</u> 27-Sep-2020	<u>Cyl. Type</u> 108	<u>Component</u> Ethanol Nitrogen	<u>Certified Concentration</u> 0.080 ± 0.002 BrAC (208 ppm) Balance
---------------------------------	-------------------------	---	---

**Certification Traceable to N.I.S.T. RGM Ethanol Standards:**

<u>Serial No.</u>	<u>Concentration</u>	<u>Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

Analytical Method: NDIR

Digitally signed by Quality Control  
 Date: 2018.10.03 10:02:04 -05:00  
 Reason: Dry gas standard certification of analysis  
 Location: Airgas USA LLC (Lab)

Approved for Release: \_\_\_\_\_

Rod Marsala

**ISO 17025:2005 A2LA accredited. Certificate Number 3082.06**



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**ALEXANDER K. STIMSON**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/6/2021

NUMBER 210092

EXPIRES 5/6/2023

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

 STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator STIMSON, ALEXANDER  
 Permit No 210092  
 Date Issued 5/6/2021 Date Expires 5/6/2023

