



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
 07/19/2021 09:46 AM

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111670	NAME OF AGENCY Vernon county Sheriff Office	DATE OF INSPECTION 07/19/2021
LOCATION OF INSTRUMENT (STREET AND CITY) 2040 E. Hunter Nevada MO.		TIME OF INSPECTION 2350

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION

COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER intoximeters LOT # AG019902 EXP. DATE 07/17/2022

SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIM. SN SIM. NIST EXP DATE

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 → .082

TEST 2 → .081

TEST 3 → .081

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0 | (0-.04) 0 | (.05-.09) 0 | (.10-.14) 1 | (.15-.19) 0 | (OVER .19) 1

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE

PRINT NAME **A. Stimson**

TYPE II PERMIT NUMBER/EXPIRATION DATE **210092 5/6/2023**

TELEPHONE NUMBER **(417-283-4400)**

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
ALEXANDER K. STIMSON

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/6/2021

NUMBER 210092

EXPIRES 5/6/2023

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator STIMSON, ALEXANDER
 Permit No 210092
 Date Issued 5/6/2021 Date Expires 5/6/2023





Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 2-Oct-2018

Lot # AG827002 Model 108cadd

<u>Exp. Date</u> 27-Sep-2020	<u>Cyl. Type</u> 108	<u>Component</u> Ethanol Nitrogen	<u>Certified Concentration</u> 0.080 ± 0.002 BrAC (208 ppm) Balance
---------------------------------	-------------------------	---	---

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

<u>Serial No.</u>	<u>Concentration</u>	<u>Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
 Date: 2018.10.03 10:02:04 -05:00
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)

Approved for Release: _____

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06

AS IV Serial no: 111670
Version no: 532B

TEST RECORD 00266

Temp Date Time 210L
%/

Air Blank: 07/19/21 23:49 .000
Calibration Check: 22 07/19/21 23:49 .082

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 111670
Version no: 532B

TEST RECORD 00267

Temp Date Time 210L
%/

Air Blank: 07/19/21 23:51 .000
Calibration Check: 23 07/19/21 23:51 .081

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 111670
Version no: 532B

TEST RECORD 00268

Temp Date Time 210L
%/

Air Blank: 07/19/21 23:52 .000
Calibration Check: 23 07/19/21 23:52 .081

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 111670
Version no: 532B

TEST RECORD 00269

Temp Date Time 210L
%/

VOID: RFI
12 07/19/21 23:54

Subject Name

Subject I.D.

Operator Name, I.D.

Location