



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
MARK D. GREEN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 9/9/2021

Laura E. Way

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 210199

Donald S. Koenig

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

EXPIRES 9/9/2023

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator GREEN, MARK
Permit No 210199
Date Issued 9/9/2021 **Date Expires** 9/9/2023





RepCo Marketing Co
3101-188 Stony Brook Drive
Raleigh, NC 27604
888-828-0227

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing Co.
LOT NUMBER: 19002
EXPIRATION DATE: October 16, 2021 at 11:59 p.m.

RepCo Marketing Co. certifies the following:


RepCo Marketing Co. prepared, tested and supplied Lot Number 19002 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by ISO 17025 and ISO 17034 accredited institution Alcohol Countermeasure Systems, using NIST standards. Random samples were analyzed by Alcohol Countermeasure Systems utilizing a gas chromatograph and found to contain .1231 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is October 17, 2019 The expiration date for this lot number is October 16, 2021 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.


Alma Palmer, Operations Manager
RepCo Marketing Co.

AS IV Serial no: 111672
Version no: 532B

TEST RECORD 00213

Temp	Date	Time	s/
22	09/28/21	12:18	

Air Blank:
09/28/21 12:18 .000

Calibration Check:
24 09/28/21 12:18 .696

Subject Name
M.D. GREEN

Subject I.D.
940

Operator Name, I.D.
M.D. Green 940

Location
Zone Office

Monthly Report

AS IV Serial no: 111672
Version no: 532B

TEST RECORD 00212

Temp	Date	Time	s/
23	09/28/21	12:17	

Air Blank:
09/28/21 12:17 .000

Calibration Check:
24 09/28/21 12:17 .696

Subject Name
M.D. GREEN

Subject I.D.
940

Operator Name, I.D.
M.D. Green 940

Location
Zone Office

Monthly Report

AS IV Serial no: 111672
Version no: 532B

TEST RECORD 00211

Temp	Date	Time	s/
23	09/28/21	12:15	

Air Blank:
09/28/21 12:15 .000

Calibration Check:
23 09/28/21 12:15 .097

Subject Name
M.D. Green

Subject I.D.
940

Operator Name, I.D.
M.D. Green 940

Location
Zone Office

Monthly Report

AS IV Serial no: 111672
Version no: 532B

TEST RECORD 00210

Temp	Date	Time	s/
23	09/28/21	12:13	

Air Blank:
09/28/21 12:13 .000

Calibration Check:
23 09/28/21 12:13 .097

Subject Name
M.D. GREEN

Subject I.D.
940

Operator Name, I.D.
M.D. Green 940

Location
Zone Office

Monthly Report