



ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111672	PRINTER SN 09B.3589.499	DATE OF INSPECTION 12/17/2021
LOCATION OF INSTRUMENT (STREET AND CITY) Zone 21 Office - Port of Kimberling Hotel - Kimberling City, MO		TIME OF INSPECTION 2:00 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Repco Marketing Inc. LOT # 20001 EXP. DATE 10/07/2022

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 33.99 SIMULATOR SN MP2307 SIMULATOR EXP DATE 01/07/2022

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .100

TEST 2 ← .099

TEST 3 ← .099

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

This instrument is operating within Dept. of Health standards. .10 Solution used.

INSPECTING OFFICER

SIGNATURE <i>Mark D. Green</i>	PRINT NAME Mark D. Green
-----------------------------------	-----------------------------

TYPE II PERMIT NUMBER/EXPIRATION DATE 210199 09-09-2023	TELEPHONE NUMBER (417) 895-6868
--	------------------------------------

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AS IV Serial no: 111672
Version no: 5328

TEST RECORD 00218

Temp Date Time 210L s/

12/17/21 16:12

Air Blank:

12/17/21 16:12 .000

Calibration Check:

20 12/17/21 16:12 .000

Subject Name

M.D. Green

Subject I.D.

940

Operator Name, I.D.

M.D. Green 940

Location

Zone Office

Mentally Maint

AS IV Serial no: 111672
Version no: 5328

TEST RECORD 00217

Temp Date Time 210L s/

12/17/21 16:11 .000

Air Blank:

12/17/21 16:11 .000

Calibration Check:

20 12/17/21 16:11 .000

Subject Name

M.D. Green

Subject I.D.

940

Operator Name, I.D.

M.D. Green 940

Location

Zone Office

Mentally Maint

AS IV Serial no: 111672
Version no: 5328

TEST RECORD 00216

Temp Date Time 210L s/

12/17/21 16:10 .000

Air Blank:

12/17/21 16:10 .000

Calibration Check:

19 12/17/21 16:10 .000

Subject Name

M.D. Green

Subject I.D.

940

Operator Name, I.D.

M.D. Green 940

Location

Zone Office

Mentally Maint

AS IV Serial no: 111672
Version no: 5328

TEST RECORD 00215

Temp Date Time 210L s/

12/17/21 16:09 .000

Air Blank:

12/17/21 16:09 .000

Calibration Check:

18 12/17/21 16:09 .100

Subject Name

M.D. Green

Subject I.D.

940

Operator Name, I.D.

M.D. Green 940

Location

Zone Office

Mentally Maint



RepCo Marketing Co
3101-188 Stony Brook Driv
Raleigh, NC 27604
919-876-5480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing Co.
LOT NUMBER: 20001
EXPIRATION DATE: October 7, 2022 at 11:59 p.m.

RepCo Marketing Co. certifies the following:

RepCo Marketing Co. prepared, tested and supplied Lot Number 20001 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by ISO 17025 and ISO 17034 accredited institution Alcohol Countermeasure Systems, using NIST standards. Random samples were analyzed by Alcohol Countermeasure Systems utilizing a gas chromatograph and found to contain .1227 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is October 8, 2020 The expiration date for this lot number is October 7, 2022 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.

Alma Palmer, Operations Manager
RepCo Marketing Co.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
MARK D. GREEN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 9/9/2021

NUMBER 210199

EXPIRES 9/9/2023

Laura E. Way

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Donald A. Kamm

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB 4 (R6-10)

 **STATE OF MISSOURI**
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator GREEN, MARK
 Permit No 210199
 Date Issued 9/9/2021 Date Expires 9/9/2023

