



ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance of the instrument.
Send copy to Department of Health and Senior Services; retain original in department file.

RECEIVED ever instrument is repaired.
By Tracy Crews at 8:30 am, Nov 12, 2021

ALCO SENSOR IV SN 111676	NAME OF AGENCY Jefferson County Sheriff's Office	DATE OF INSPECTION 11/10/2021
LOCATION OF INSTRUMENT (STREET AND CITY) 410 First Street, Hillsboro, MO 63050		TIME OF INSPECTION 6:42 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION

COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Intoximeter LOT # AG109702 EXP. DATE 04/07/2023

SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIM. SN SIM. NIST EXP DATE

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ➔ .103

TEST 2 ➔ .102

TEST 3 ➔ .102

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	---------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE

 549

PRINT NAME

Nicholas Gamm

TYPE II PERMIT NUMBER/EXPIRATION DATE

200190 / 06/26/2022

TELEPHONE NUMBER

(636) 797-5000

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111676
Version no: 532B

TEST RECORD 00172

Temp	Date	Time	^{a/} 210L
------	------	------	--------------------

Air Blank:
11/18/21 19:25 .000
Calibration Check:
20 11/18/21 19:25 .103

Subject Name

Nov Maintenance

Subject I.D.

Gamm 200190

Operator Name, I.D.

Exp: 06/26/2022

Location

JCSO HQ

Test #1

AS IV Serial no: 111676
Version no: 532B

TEST RECORD 00173

Temp	Date	Time	^{a/} 210L
------	------	------	--------------------

Air Blank:
11/18/21 19:28 .000
Calibration Check:
21 11/18/21 19:28 .102

Subject Name

Subject I.D.

Operator Name, I.D.

Location

Test #2

AS IV Serial no: 111676
Version no: 532B

TEST RECORD 00174

Temp	Date	Time	^{a/} 210L
------	------	------	--------------------

Air Blank:
11/18/21 19:31 .000
Calibration Check:
22 11/18/21 19:31 .102

Subject Name

Subject I.D.

Operator Name, I.D.

Location

Test #3

AS IV Serial no: 111676
Version no: 532B

TEST RECORD 00175

Temp	Date	Time	^{a/} 210L
------	------	------	--------------------

VOID: RFI
12 11/18/21 19:33

Subject Name

Subject I.D.

Operator Name, I.D.

Location

RFI!!!



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 8-Apr-2021

Lot # AG109702 Model 108caccd

<u>Exp. Date</u> 7-Apr-2023	<u>Cyl. Type</u> 108	<u>Component</u> Ethanol Nitrogen	<u>Certified Concentration</u> 0.100 ± 2% BrAC (272 ppm) Balance
--------------------------------	-------------------------	---	--

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
 Date: 2021.04.08 18:24:39 -05:00
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)

Approved for Release: _____

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07