

AS IV Serial no: 111677
Version no: 532B

TEST RECORD 00643

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
02/28/21 14:48 .000

Calibration Check:
24 02/28/21 14:48 .096

Subject Name

TEST 1

Subject I.D.

Operator Name, I.D.

E.J. GADJME 200019
Location

MADISON CO. JAIL

AS IV Serial no: 111677
Version no: 532B

TEST RECORD 00644

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
02/28/21 14:50 .000

Calibration Check:
24 02/28/21 14:50 .096

Subject Name

TEST 2

Subject I.D.

Operator Name, I.D.

E.J. GADJME 200019
Location

MADISON CO. JAIL

AS IV Serial no: 111677
Version no: 532B

TEST RECORD 00645

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
02/28/21 14:51 .000

Calibration Check:
25 02/28/21 14:51 .096

Subject Name

TEST 3

Subject I.D.

Operator Name, I.D.

E.J. GADJME 200019
Location

MADISON CO. JAIL

AS IV Serial no: 111677
Version no: 532B

TEST RECORD 00646

Temp	Date	Time	s/ 210L
------	------	------	------------

VOID: RFI
12 02/28/21 14:52

Subject Name

RFI

Subject I.D.

Operator Name, I.D.

E.J. GADJME 200019
Location

MADISON CO. JAIL



RepCo Marketing Co
3101-188 Stony Brook Drive
Raleigh, NC 27604
888-828-0227

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing Co.

LOT NUMBER: 19001

EXPIRATION DATE: March 12, 2021 at 11:59 p.m.

RepCo Marketing Co. certifies the following:

RepCo Marketing Co. prepared, tested and supplied Lot Number 19001 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by accredited institution, Data Resources Inc., using NIST standards. Random samples were analyzed by Data Resources Inc. utilizing a gas chromatograph and found to contain .1210 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 13, 2019
The expiration date for this lot number is March 12, 2021 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.

Alma Palmer, Operations Manager
RepCo Marketing Co.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
ERIC J GANIME

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 1/6/2020

NUMBER 200019

EXPIRES 1/6/2022

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator GANIME, ERIC
Permit No 200019
Date Issued 1/6/2020 **Date Expires** 1/6/2022