

RECEIVED

By Tracy Crews at 11:47 am, Feb 17, 2021

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111732	PRINTER SN 09B.3589.454	DATE OF INSPECTION 02/14/2021
LOCATION OF INSTRUMENT (STREET AND CITY) 201 E Brick Street Ozark Mo 65721		TIME OF INSPECTION 8:09 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument. DIGITAL READOUT (ALL ELEMENTS OPERATIONAL) TEMPERATURE OF ALCO SENSOR (10°C - 40°C) PRINTER WORKING PROPERLY TIME AND DATE DISPLAYING PROPERLY**BREATH ALCOHOL ACCURACY STANDARDS** SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE STANDARD SUPPLIER Guth LOT # 19370 EXP. DATE 12/9/2021 SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN DR6932 SIMULATOR EXP DATE 12/8/2021 CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .098

TEST 2 .096

TEST 3 .096

 RFI DETECTOR OPERATING**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS 0 | (0-.04) 0 | (.05-.09) 0 | (.10-.14) 0 | (.15-.19) 0 | (OVER .19) 0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE

PRINT NAME

Justin Brown

TYPE II PERMIT NUMBER/EXPIRATION DATE

TELEPHONE NUMBER

417-581-6600

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AS IV Serial no: 111732
Version no: 532B

TEST RECORD 00737

Temp Date Time ^{s/} 210L

Air Blank:
02/14/21 20:09 .000
Calibration Check:
24 02/14/21 20:09 .098

Subject Name

Test 1

Subject I.D.

Operator Name, I.D.

J. Brown / 200297

Location

201 E Brick Street

Ozark, Mo, 65721

AS IV Serial no: 111732
Version no: 532B

TEST RECORD 00738

Temp Date Time ^{s/} 210L

Air Blank:
02/14/21 20:11 .000
Calibration Check:
25 02/14/21 20:11 .096

Subject Name

Test 2

Subject I.D.

Operator Name, I.D.

J. Brown / 200297

Location

201 E Brick Street

Ozark, Mo, 65721

AS IV Serial no: 111732
Version no: 532B

TEST RECORD 00739

Temp Date Time ^{s/} 210L

Air Blank:
02/14/21 20:13 .000
Calibration Check:
25 02/14/21 20:13 .096

Subject Name

Test 3

Subject I.D.

Operator Name, I.D.

J. Brown / 200297

Location

201 E Brick Street

Ozark, Mo, 65721

AS IV Serial no: 111732
Version no: 532B

TEST RECORD 00740

Temp Date Time ^{s/} 210L

VOID: RFI
12 02/14/21 20:15

Subject Name

RFI

Subject I.D.

Operator Name, I.D.

J. Brown / 200297

Location

201 E Brick Street

Ozark, Mo, 65721



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4811 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 19370 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **December 10, 2019**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1199%** (w/vol) ethyl alcohol. The expiration date for this lot number is **December 9, 2021** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

*Testing was conducted using Cerilliant Reference Standard lot number FN04271602 whose values are traceable to NIST.
All balances are calibrated annually by an outside agency using NIST traceable weights.
Calibration verification is done prior to each use utilizing NIST traceable weights.*



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
JUSTIN BROWN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/11/2020

NUMBER 200297

EXPIRES 12/11/2022

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (5-10)

LAB-4 (R5-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator BROWN, JUSTIN
Permit No 200297
Date Issued 12/11/2020 **Date Expires** 12/11/2022

