



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111740 Printer #09B.3589.511	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 02/18/2021
LOCATION OF INSTRUMENT (STREET AND CITY) Missouri Department of Transportation, 1540 N. Highway 63, Houston, Missouri 65483		TIME OF INSPECTION 9:50 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER RepCo Marketing Co. LOT # 19002 EXP. DATE 10/16/2021

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 33.96 SIM. SN MP2124 SIM. NIST EXP DATE 02/03/2022

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .095

TEST 2 ← .096

TEST 3 ← .095

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE *Dale B Pounds*

PRINT NAME
Dale B Pounds

TYPE II PERMIT NUMBER/EXPIRATION DATE
200214 07/21/2022

TELEPHONE NUMBER
(417) 469-3121

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111740
Version no: 532B

TEST RECORD 00067

Temp Date Time ^{g/} 210L

Air Blank:
02/18/21 12:40 .000
Calibration Check:
23 02/18/21 12:40 .095

Subject Name

TEST #1

Subject I.D.

MONTHLY MAINTENANCE

Operator Name, I.D.

Pounds, D.B. #200214

Location

MISSOURI DEPT TRANSPORTATION

1540 N. HIGHWAY 63

HOUSTON, MISSOURI 65483

AS IV Serial no: 111740
Version no: 532B

TEST RECORD 00068

Temp Date Time ^{g/} 210L

Air Blank:
02/18/21 12:42 .000
Calibration Check:
24 02/18/21 12:42 .096

Subject Name

TEST #2

Subject I.D.

MONTHLY MAINTENANCE REPORT

Operator Name, I.D.

Pounds, D.B. #200214

MISSOURI DEPARTMENT OF TRANSPORTATION

Location

1540 N. HIGHWAY 63

HOUSTON, MISSOURI 65483

AS IV Serial no: 111740
Version no: 532B

TEST RECORD 00069

Temp Date Time ^{g/} 210L

Air Blank:
02/18/21 12:43 .000
Calibration Check:
24 02/18/21 12:43 .095

Subject Name

TEST #3

Subject I.D.

MONTHLY MAINTENANCE REPORT

Operator Name, I.D.

Pounds, D.B. #200214

Location

MISSOURI DEPARTMENT OF TRANSPORTATION

1540 N. HIGHWAY 63

HOUSTON, MISSOURI 65483

AS IV Serial no: 111740
Version no: 532B

TEST RECORD 00070

Temp Date Time ^{g/} 210L

VOID: RFI
12 02/18/21 12:45

Subject Name

TEST #4

Subject I.D.

MONTHLY MAINTENANCE

Operator Name, I.D.

Pounds, D.B. #200214

Location

MISSOURI DEPARTMENT OF TRANSPORTATION

1540 N. HIGHWAY 63

HOUSTON, MISSOURI 65483

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing Co.

LOT NUMBER: 19002

EXPIRATION DATE: October 16, 2021 at 11:59 p.m.

RepCo Marketing Co. certifies the following:

RepCo Marketing Co. prepared, tested and supplied Lot Number 19002 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by ISO 17025 and ISO 17034 accredited institution Alcohol Countermeasure Systems, using NIST standards. Random samples were analyzed by Alcohol Countermeasure Systems utilizing a gas chromatograph and found to contain .1231 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is October 17, 2019 The expiration date for this lot number is October 16, 2021 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Alma Palmer, Operations Manager
RepCo Marketing Co.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
DALE B POUNDS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/21/2020

NUMBER 200214

EXPIRES 7/21/2022

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator POUNDS, DALE
 Permit No 200214
 Date Issued 7/21/2020 Date Expires 7/21/2022