



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY

**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111751	NAME OF AGENCY Boone County Sheriff's Department	DATE OF INSPECTION 01/12/2021
LOCATION OF INSTRUMENT (STREET AND CITY) 2111 E County Dr, Columbia, MO 65202		TIME OF INSPECTION 4:07 am

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

<input type="checkbox"/> SIMULATOR SOLUTION	<input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>Intoximeters</u>	LOT # <u>AG023902</u> EXP. DATE <u>08/26/2022</u>
<input type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____	SIM. SN _____ SIM. NIST EXP DATE _____

**CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**  
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1  .099	TEST 2  .100	TEST 3  .098
--------------	--------------	--------------

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	---------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER	
SIGNATURE 	PRINT NAME Brandon Baker
TYPE II PERMIT NUMBER/EXPIRATION DATE 200250, 09/24/2022	TELEPHONE NUMBER 573-875-1111

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111751  
Version no: 532B

TEST RECORD 00367

Temp Date Time 210L  
9/

Air Blank:  
01/12/21 04:57 .000  
Calibration Check:  
23 01/12/21 04:57 .099

Subject Name

TEST 1

Subject I.D.

Operator Name, I.D.

BRANDON BAYLER

Location 200350

2111 E COURT Y DR

AS IV Serial no: 111751  
Version no: 532B

TEST RECORD 00368

Temp Date Time 210L  
9/

Air Blank:  
01/12/21 04:59 .000  
Calibration Check:  
23 01/12/21 04:59 .100

Subject Name

TEST 2

Subject I.D.

Operator Name, I.D.

BRANDON BAYLER

Location 200350

2111 E COURT Y DR

AS IV Serial no: 111751  
Version no: 532B

TEST RECORD 00369

Temp Date Time 210L  
9/

Air Blank:  
01/12/21 05:00 .000  
Calibration Check:  
24 01/12/21 05:00 .098

Subject Name

TEST 3

Subject I.D.

Operator Name, I.D.

BRANDON BAYLER

Location 200350

2111 E COURT Y DR

AS IV Serial no: 111751  
Version no: 532B

TEST RECORD 00370

Temp Date Time 210L  
9/

VOID: RFI  
12 01/12/21 05:02

Subject Name

RFI

Subject I.D.

Operator Name, I.D.

BRANDON BAYLER

Location 200350

2111 E COURT Y DR

Subject Name  
Blank

Subject I.D.

Operator Name, I.D.

BRANDON BAYLER

Location 200350

2111 E COURT Y DR

AS IV Serial no: 111751  
Version no: 532B  
TEST RECORD 00371  
Temp Date Time 210L  
9/  
Air Blank:  
01/12/21 05:04 .000  
Subject Test: Auto  
24 01/12/21 05:04 .000



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**

*Exclusive Supplier*

Intoximeters, Inc.

2081 Craig Road

St. Louis, Mo 63146

**Test Date:** 27-Aug-2020

**Lot # AG023902 Model 108cacd**

**Exp. Date**

26-Aug-2022

**Cyl. Type**

108

**Component**

Ethanol ✓

Nitrogen

**Certified Concentration**

0.100 ± 2% BrAC (272 ppm)

Balance

**Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:**

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm
<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC434668	800.0 ppm	0056649	390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

**Analytical Method:** NDIR

Digitally signed by Quality Control  
Date: 2020.08.27 19:06:48 -05:00  
Reason: Dry gas standard certification of analysis  
Location: Airgas USA LLC (Lab)

Approved for Release: \_\_\_\_\_

Rod Marsala

**ISO 17025:2005 A2LA accredited. Certificate Number 3082.06**  
**ISO 17034:2016 A2LA accredited. Certificate Number 3082.07**



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM



**PERMIT  
 TYPE II**

**BRANDON BAKER**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX DMT**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 9/24/2020

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 200250

EXPIRES 9/24/2022

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 586-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator BAKER, BRANDON  
 Permit No 200250  
 Date Issued 9/24/2020 Date Expires 9/24/2022