

RECEIVED

By Tracy Crews at 9:25 am, Mar 10, 2021

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111751	NAME OF AGENCY Boone County Sheriff's Office	DATE OF INSPECTION 03/10/2021
LOCATION OF INSTRUMENT (STREET AND CITY) 2111 E County Drive, Columbia, MO 65202		TIME OF INSPECTION 5:03 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument. DIGITAL READOUT (ALL ELEMENTS OPERATIONAL) TEMPERATURE OF ALCO SENSOR (10°C - 40°C) PRINTER WORKING PROPERLY TIME AND DATE DISPLAYING PROPERLY**BREATH ALCOHOL ACCURACY STANDARDS** SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE STANDARD SUPPLIER Intoximeters LOT # AG023902 EXP. DATE 08/26/2022 SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIM. SN SIM. NIST EXP DATE CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .096

TEST 2 .096

TEST 3 .096

 RFI DETECTOR OPERATING**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS

(0-.04)

(.05-.09)

(.10-.14)

(.15-.19)

(OVER .19)

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE

PRINT NAME

Brandon Baker

TYPE II PERMIT NUMBER/EXPIRATION DATE

200250 09/24/2022

TELEPHONE NUMBER

(573) 875-1111

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111751
Version no: 532B

TEST RECORD 00391

Temp Date Time 210L
Air Blank: 03/10/21 05:03 .000
Calibration Check: 26 03/10/21 05:03 .096

Subject Name
TEST 1
Subject I.D.

Operator Name, I.D.
BRANDON BAWER
Location 200350
2111 E COUNTY DR

AS IV Serial no: 111751
Version no: 532B

TEST RECORD 00392

Temp Date Time 210L
Air Blank: 03/10/21 05:04 .000
Calibration Check: 26 03/10/21 05:04 .096

Subject Name
TEST 2
Subject I.D.

Operator Name, I.D.
BRANDON BAWER
Location 200350
2111 E COUNTY DR

AS IV Serial no: 111751
Version no: 532B

TEST RECORD 00393

Temp Date Time 210L
Air Blank: 03/10/21 05:06 .000
Calibration Check: 27 03/10/21 05:06 .096

Subject Name
TEST 3
Subject I.D.

Operator Name, I.D.
BRANDON BAWER
Location 200350
2111 E COUNTY DR

AS IV Serial no: 111751
Version no: 532B

TEST RECORD 00394

Temp Date Time 210L
VOID: RFI
12 03/10/21 05:07

Subject Name
RFI
Subject I.D.

Operator Name, I.D.
BRANDON BAWER
Location 200350
2111 E COUNTY DR



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
BRANDON BAKER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 9/24/2020

NUMBER 200250

EXPIRES 9/24/2022

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

 **STATE OF MISSOURI**
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator BAKER, BRANDON
Permit No 200250
Date Issued 9/24/2020 **Date Expires** 9/24/2022





Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

Certificate of Analysis

Customer Name

Exclusive Supplier

Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 27-Aug-2020

Lot # AG023902 **Model** 108cacc

Exp. Date

26-Aug-2022

Cyl. Type

108

Component

Ethanol
Nitrogen

Certified Concentration

0.100 ± 2% BrAC (272 ppm)
Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm
<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC434668	800.0 ppm	0056649	390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
Date: 2020.08.27 19:06:48 -05:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Approved for Release:

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07