

AS IU Serial no: 111753
Version no: 532B

TEST RECORD 00322

Temp Date Time 210L
9/

Air Blank:
01/06/21 23:17 .000
Calibration Check:
20 01/06/21 23:17 .101

Subject Name

Test 1 00322

Subject I.D.

Hutchinson 171

Operator Name, I.D.

200128

Location

Greenwood PD

AS IU Serial no: 111753
Version no: 532B

TEST RECORD 00323

Temp Date Time 210L
9/

Air Blank:
01/06/21 23:21 .000
Calibration Check:
21 01/06/21 23:21 .101

Subject Name

Test 2 06323

Subject I.D.

Hutchinson 171

Operator Name, I.D.

200128

Location

Greenwood PD

AS IU Serial no: 111753
Version no: 532B

TEST RECORD 00324

Temp Date Time 210L
9/

Air Blank:
01/06/21 23:23 .000
Calibration Check:
21 01/06/21 23:23 .101

Subject Name

Test 3 00324

Subject I.D.

Hutchinson 171

Operator Name, I.D.

200128

Location

Greenwood PD

AS IU Serial no: 111753
Version no: 532B

TEST RECORD 00325

Temp Date Time 210L
9/

VOID: RFI
12 01/06/21 23:26

Subject Name

Test 4 0325

Subject I.D.

Hutchinson 171

Operator Name, I.D.

200128

Location

Greenwood PD

Monthly AS-IV Maintenance Check

Greenwood PD 1-6-2021



GUTH LABORATORIES, INC.

890 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-664-8470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 19370 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on December 10, 2019, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1199% (w/vol) ethyl alcohol. The expiration date for this lot number is December 9, 2021 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN04271602 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

**PERMIT
 TYPE II**

ANDREW HUTCHINSON

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/9/2020

NUMBER 200128

EXPIRES 3/9/2022

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator HUTCHINSON, ANDREW
Permit No 200128
Date Issued 3/9/2020 **Date Expires** 3/9/2022

